

Analysis Method: 9050A

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124729

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP101736

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB124729BL		1	24.2	0.82	0.83	03/28/2023	16:00
2	LB124729BS	1413	1	23.0	1370.00	1424.41	03/28/2023	16:05
3	O1953-09		1	24.7	552.00	555.18	03/28/2023	16:10

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$