



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124730

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/29/2023	08:25
2	CAL2	1	Water	NA	NA	20.7	7.00	03/29/2023	08:26
3	CAL3	1	Water	NA	NA	20.7	10.03	03/29/2023	08:29
4	ICV	1	Water	NA	NA	20.7	6.98	03/29/2023	08:30
5	CCV1	1	Water	NA	NA	20.7	2.01	03/29/2023	08:33
6	O2110-01	1	Water	NA	NA	20.2	9.17	03/29/2023	08:40
7	O2110-01DUP	1	Water	NA	NA	20.1	9.18	03/29/2023	08:41
8	CCV2	1	Water	NA	NA	20.7	12.02	03/29/2023	08:45

WORKLIST(Hardcopy Internal Chain)

WB 124430

WorkList Name : ph o2110

WorkList ID : 168575

Department : Wet-Chemistry

Date : 03-29-2023 08:16:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/27/2023	Water	O2110-01	pH		Cool 4 deg C	PSEG03 J11	N28230	03/27/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

03/29/23 08:20

JP (WC)

JP

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

03/29/23

9:30

JP

JP (WC)