



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124735

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/29/2023	12:25
2	CAL2	1	Water	NA	NA	20.7	7.01	03/29/2023	12:26
3	CAL3	1	Water	NA	NA	20.7	10.03	03/29/2023	12:27
4	ICV	1	Water	NA	NA	20.6	7.00	03/29/2023	12:30
5	CCV1	1	Water	NA	NA	20.8	2.01	03/29/2023	12:33
6	O2124-01	1	Water	NA	NA	20.2	6.50	03/29/2023	12:38
7	O2124-01DUP	1	Water	NA	NA	20.3	6.51	03/29/2023	12:39
8	CCV2	1	Water	NA	NA	20.7	12.03	03/29/2023	12:40

2031241235

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 02124

WorkList ID : 168598

Department : Wet-Chemistry

Date : 03-29-2023 12:07:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/29/2023	Water	02124-01	pH	D	Cool 4 deg C	ENTA05	M11	E9074-PVSC-001	03/29/2023 9040C

Date/Time 03/29/23 12:15
Raw Sample Received by: 2P/WCJ
Raw Sample Relinquished by: 2P/WCJ

Date/Time 03/29/23 12:130
Raw Sample Received by: 2P/WCJ
Raw Sample Relinquished by: 2P/WCJ