



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124742

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	03/29/2023	08:50
2	O2114-01		1	14	100.00	>212.0	>212.0	03/29/2023	09:20
3	O2114-01DUP		1	14	100.00	>212.0	>212.0	03/29/2023	09:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6124742

WorkList Name : f p s 3-29

WorkList ID : 168571

Department : Wet-Chemistry

Date : 03-28-2023 16:44:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/04/2023	Solid	O2114-01	Flash Point	Cool 4 deg C	PSEG03	K11	15091	03/28/2023	1010B

Date/Time 03/29/2023 08:40
 Raw Sample Received by: RMC
 Raw Sample Relinquished by: JPC

Date/Time 03/29/2023 10:36
 Raw Sample Received by: JPC
 Raw Sample Relinquished by: RMC