


LB124743 INSTRUMENT: CVI

7470A

Linear

μ Abs.:

40306



A= 0.0000e+000

B= 2.4738e-004

C= -8.2447e-002

Rho= 0.9991203

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.014	0.014	391	0.000	391	0				0
0.2	0.200	0.240	0.040	1304	0.0 %	1304					20
2.5	2.500	2.507	0.007	10467	0.0 %	10467					0
5.0	5.000	4.772	-0.228	19624	0.0 %	19624					-5
7.5	7.500	7.778	0.278	31776	0.0 %	31776					4
10.0	10.000	9.888	-0.112	40306	0.0 %	40306					1

LB124743 INSTRUMENT ID: CV1

Method: 7470A		Operator: Admin		Date of Analysis: 29 Mar 2023 16:05:21					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	391	-	0.0000	7470A	PPB	29 Mar 2023 16:22:57	S	Std
0.2 - 1	50.2	1304	-	0.2000	7470A	PPB	29 Mar 2023 16:25:14	S	Std
2.5 - 1	52.5	10467	-	2.5000	7470A	PPB	29 Mar 2023 16:27:31	S	Std
5.0 - 1	55.0	19624	-	5.0000	7470A	PPB	29 Mar 2023 16:29:48	S	Std
7.5 - 1	57.5	31776	-	7.5000	7470A	PPB	29 Mar 2023 16:32:05	S	Std
10.0 - 1	59.0	40306	-	10.0000	7470A	PPB	29 Mar 2023 16:36:53	S	Std
ICV04 - 1	ICV04	17930	4.3530	-	7470A	PPB	29 Mar 2023 16:39:49	U	SMPL
ICB04 - 1	ICB04	287	-0.0114	-	7470A	PPB	29 Mar 2023 16:42:04	U	SMPL
CCV15 - 1	CCV15	20135	4.8985	-	7470A	PPB	29 Mar 2023 16:44:20	U	SMPL
CCB15 - 1	CCB15	258	-0.0186	-	7470A	PPB	29 Mar 2023 16:46:36	U	SMPL
CRA - 1	CRA	1228	0.2213	-	7470A	PPB	29 Mar 2023 16:48:52	U	SMPL
HighStd - 1	HighStd	37731	9.2514	-	7470A	PPB	29 Mar 2023 16:51:07	U	SMPL
ChkStd - 1	ChkStd	26367	6.4402	-	7470A	PPB	29 Mar 2023 16:53:23	U	SMPL
PB151773BL - 1	PBW	224	-0.0270	-	7470A	PPB	29 Mar 2023 16:55:40	U	SMPL
PB151773BS - 1	LCSW	16550	4.0117	-	7470A	PPB	29 Mar 2023 16:57:57	U	SMPL
O2069-01 - 1	API-3-3	516	0.0452	-	7470A	PPB	29 Mar 2023 17:00:14	U	SMPL
O2069-02 - 1	API-3-4	574	0.0595	-	7470A	PPB	29 Mar 2023 17:02:31	U	SMPL
O2104-03 - 1	EB-2023-03-25	299	-0.0085	-	7470A	PPB	29 Mar 2023 17:04:48	U	SMPL
O2104-05 - 1	B-37-GW01	31829	7.7914	-	7470A	PPB	29 Mar 2023 17:07:05	U	SMPL
O2109-03 - 1	B-30-GW01	-35	-0.0911	-	7470A	PPB	29 Mar 2023 17:09:21	U	SMPL
CCV16 - 1	CCV16	21354	5.2001	-	7470A	PPB	29 Mar 2023 17:16:21	U	SMPL
CCB16 - 1	CCB16	146	-0.0463	-	7470A	PPB	29 Mar 2023 17:18:36	U	SMPL
O2109-04 - 1	EB-2023-03-26	266	-0.0166	-	7470A	PPB	29 Mar 2023 17:20:53	U	SMPL
O2110-01 - 1	N28230	292	-0.0102	-	7470A	PPB	29 Mar 2023 17:23:08	U	SMPL
O2110-01DUP - 1	N28230DUP	303	-0.0075	-	7470A	PPB	29 Mar 2023 17:25:24	U	SMPL
O2110-01MS - 1	N28230MS	16864	4.0893	-	7470A	PPB	29 Mar 2023 17:32:30	U	SMPL
O2110-01MSD - 1	N28230MSD	17231	4.1801	-	7470A	PPB	29 Mar 2023 17:34:46	U	SMPL
PB151791BL - 1	PBW	390	0.0140	-	7470A	PPB	29 Mar 2023 17:37:04	U	SMPL
PB151791BS - 1	LCSW	16355	3.9634	-	7470A	PPB	29 Mar 2023 17:39:22	U	SMPL
O2102-02 - 1	CHAMBER-C-72-EAST	93	-0.0594	-	7470A	PPB	29 Mar 2023 17:50:17	U	SMPL
O2105-02 - 1	WV-SOIL1-0327	100	-0.0577	-	7470A	PPB	29 Mar 2023 17:52:34	U	SMPL
O2105-05 - 1	WV-SOIL2-0327	112	-0.0547	-	7470A	PPB	29 Mar 2023 17:54:49	U	SMPL
CCV17 - 1	CCV17	20502	4.9893	-	7470A	PPB	29 Mar 2023 18:01:18	U	SMPL
CCB17 - 1	CCB17	132	-0.0498	-	7470A	PPB	29 Mar 2023 18:03:33	U	SMPL
O2105-08 - 1	WV-SOIL3-0327	122	-0.0523	-	7470A	PPB	29 Mar 2023 18:05:50	U	SMPL
O2107-02 - 1	WC01	133	-0.0495	-	7470A	PPB	29 Mar 2023 18:08:06	U	SMPL
O2107-02DUP - 1	WC01DUP	138	-0.0483	-	7470A	PPB	29 Mar 2023 18:10:21	U	SMPL
O2107-02MS - 1	WC01MS	12736	3.0682	-	7470A	PPB	29 Mar 2023 18:12:37	U	SMPL
O2107-02MSD - 1	WC01MSD	14249	3.4424	-	7470A	PPB	29 Mar 2023 18:14:53	U	SMPL
O2110-01LX5 - 1		203	-0.0322	-	7470A	PPB	29 Mar 2023 18:17:11	U	SMPL
O2110-01A - 1		17438	4.2313	-	7470A	PPB	29 Mar 2023 18:31:57	U	SMPL
O2107-02LX5 - 1		123	-0.0520	-	7470A	PPB	29 Mar 2023 18:34:14	U	SMPL
O2107-02A - 1		16083	3.8961	-	7470A	PPB	29 Mar 2023 18:36:32	U	SMPL
PB151735TB - 1		71	-0.0649	-	7470A	PPB	29 Mar 2023 18:38:48	U	SMPL
CCV18 - 1	CCV18	19334	4.7004	-	7470A	PPB	29 Mar 2023 18:41:05	U	SMPL
CCB18 - 1	CCB18	112	-0.0547	-	7470A	PPB	29 Mar 2023 18:43:21	U	SMPL