

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

3/29/2023 15:01

Reviewed by : 12 Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.980	0.0	0.115	
ICB1	0.007	0.0	0.008	
CCV1	0.968	0.0	0.114	
CCB1	0.006	0.0	0.008	
PB151788BL	0.008	0.0	0.008	
PB151788BS	0.979	0.0	0.115	
O1953-25	0.790	0.0	0.094	
CCV2	0.956	0.0	0.113	
CCB2	0.007	0.0	0.008	
CCB2	0.003	0.0	0.008	

N 10
Mean 0.470
SD 0.4923
CV% 104.66

Aquakem v. 7.2AQ1

Results from time period:

Wed Mar 29 14:17:45 2023

Wed Mar 29 14:58:43 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0067	mg/l	3/29/2023 14:17:45	
0.05PPM	A	Phenolics-	P	0.0583	mg/l	3/29/2023 14:17:46	
0.1PPM	A	Phenolics-	P	0.1059	mg/l	3/29/2023 14:17:47	
0.25PPM	A	Phenolics-	P	0.2608	mg/l	3/29/2023 14:17:48	
0.50PPM	A	Phenolics-	P	0.4897	mg/l	3/29/2023 14:17:49	
1.0PPM	A	Phenolics-	P	0.956	mg/l	3/29/2023 14:17:50	
2.0PPM	A	Phenolics-	P	2.0227	mg/l	3/29/2023 14:17:51	
ICV1	S	Phenolics-	P	0.9799	mg/l	3/29/2023 14:51:50	
ICB1	S	Phenolics-	P	0.0074	mg/l	3/29/2023 14:51:53	
CCV1	S	Phenolics-	P	0.9684	mg/l	3/29/2023 14:51:54	
CCB1	S	Phenolics-	P	0.0061	mg/l	3/29/2023 14:51:57	
PB151788BL	S	Phenolics-	P	0.0076	mg/l	3/29/2023 14:51:59	
PB151788BS	S	Phenolics-	P	0.9788	mg/l	3/29/2023 14:52:01	
O1953-25	S	Phenolics-	P	0.7899	mg/l	3/29/2023 14:58:37	
CCV2	S	Phenolics-	P	0.9558	mg/l	3/29/2023 14:58:41	
CCB2	S	Phenolics-	P	0.0069	mg/l	3/29/2023 14:58:42	
CCB2	S	Phenolics-	P	0.0029	mg/l	3/29/2023 14:58:43	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

3/29/2023 14:21

Reviewed by : 12 Instrument ID : Konelab


Test Phenolics-

Accepted 3/29/2023 14:21

Factor 9.056
Bias 0.007

Coeff. of det. 0.999105

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.00PPM	0.008	0.0067	0.0000	
2	Phenol-2	0.014	0.0583	0.0500	16.6
3	Phenol-2	0.019	0.1059	0.1000	5.9
4	Phenol-2	0.036	0.2608	0.2500	4.3
5	Phenol-2	0.061	0.4897	0.5000	-2.1
6	Phenol-2	0.113	0.9560	1.0000	-4.4
7	Phenol-2	0.231	2.0227	2.0000	1.1

12
3/29/23