

L13124745

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

3/29/2023 15:33 Reviewed by : 12 Instrument ID : Konelab

Test: Phenolics-

Sample Id	^{mol} Result	Dil. 1 +	Response	Errors
ICV1	0.981	0.0	0.116	
ICB1	0.010	0.0	0.008	
CCV1	0.959	0.0	0.113	
CCB1	0.007	0.0	0.008	
PB151789BL	0.008	0.0	0.008	
PB151789BS	0.981	0.0	0.116	
O1953-26	0.796	0.0	0.095	
CCV2	0.961	0.0	0.113	
CCB2	0.008	0.0	0.008	

N	9
Mean	0.523
SD	0.4919
CV%	94.00

Aquakem v. 7.2AQ1

Results from time period:

Wed Mar 29 15:22:52 2023

Wed Mar 29 15:28:47 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0067	mg/l	3/29/2023 14:17:45	
0.05PPM	A	Phenolics-	P	0.0583	mg/l	3/29/2023 14:17:46	
0.1PPM	A	Phenolics-	P	0.1059	mg/l	3/29/2023 14:17:47	
0.25PPM	A	Phenolics-	P	0.2608	mg/l	3/29/2023 14:17:48	
0.50PPM	A	Phenolics-	P	0.4897	mg/l	3/29/2023 14:17:49	
1.0PPM	A	Phenolics-	P	0.956	mg/l	3/29/2023 14:17:50	
2.0PPM	A	Phenolics-	P	2.0227	mg/l	3/29/2023 14:17:51	
ICV1	S	Phenolics-	P	0.9808	mg/l	3/29/2023 15:22:53	
ICB1	S	Phenolics-	P	0.01	mg/l	3/29/2023 15:22:54	
CCV1	S	Phenolics-	P	0.959	mg/l	3/29/2023 15:22:57	
CCB1	S	Phenolics-	P	0.0069	mg/l	3/29/2023 15:22:58	
PB151789BL	S	Phenolics-	P	0.008	mg/l	3/29/2023 15:23:00	
PB151789BS	S	Phenolics-	P	0.981	mg/l	3/29/2023 15:23:03	
O1953-26	S	Phenolics-	P	0.796	mg/l	3/29/2023 15:28:42	
CCV2	S	Phenolics-	P	0.9607	mg/l	3/29/2023 15:28:45	
CCB2	S	Phenolics-	P	0.0075	mg/l	3/29/2023 15:28:47	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

3/29/2023 14:21

Reviewed by : 12 Instrument ID : Konelab


Test Phenolics-

Accepted 3/29/2023 14:21

Factor 9.056
Bias 0.007

Coeff. of det. 0.999105

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.008	0.0067	0.0000	
2	Phenol-2	0.014	0.0583	0.0500	16.6
3	Phenol-2	0.019	0.1059	0.1000	5.9
4	Phenol-2	0.036	0.2608	0.2500	4.3
5	Phenol-2	0.061	0.4897	0.5000	-2.1
6	Phenol-2	0.113	0.9560	1.0000	-4.4
7	Phenol-2	0.231	2.0227	2.0000	1.1

12
3/29/23