


LB124747 INSTRUMENT 1.CV1.

7471B

Linear

 μ Abs.:

34226



A= 0.0000e+000

B= 2.9470e-004

C= -4.1809e-002

Rho= 0.9999194

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	χ^2
0.0	0.000	0.035	0.035	259	0.000	259					19
0.2	0.200	0.238	0.038	948	0.0 %	948					-2
2.5	2.500	2.460	-0.040	8489	0.0 %	8489					-2
5.0	5.000	4.917	-0.083	16826	0.0 %	16826					0
7.5	7.500	7.507	0.007	25614	0.0 %	25614					0.5
10.0	10.000	10.045	0.045	34226	0.0 %	34226					

LB124747, INSTRUMENT ID: CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Mar 2023 19:12:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	259	-	0.0000	7471B	PPB	29 Mar 2023 19:15:52	S	Std
0.2 - 1	30.2	948	-	0.2000	7471B	PPB	29 Mar 2023 19:18:09	S	Std
2.5 - 1	52.5	8489	-	2.5000	7471B	PPB	29 Mar 2023 19:20:27	S	Std
5.0 - 1	55.0	16826	-	5.0000	7471B	PPB	29 Mar 2023 19:22:44	S	Std
7.5 - 1	57.5	25614	-	7.5000	7471B	PPB	29 Mar 2023 19:25:01	S	Std
10.0 - 1	510.0	34226	-	10.0000	7471B	PPB	29 Mar 2023 19:27:19	S	Std
ICV05 - 1	ICV058	14471	4.2228	-	7471B	PPB	29 Mar 2023 19:42:34	U	SMPL
ICB05 - 1	ICB05	-9	-0.0445	-	7471B	PPB	29 Mar 2023 19:44:49	U	SMPL
CCV19 - 1	CCV19	17941	5.2454	-	7471B	PPB	29 Mar 2023 19:47:06	U	SMPL
CCB19 - 1	CCB19	32	-0.0324	-	7471B	PPB	29 Mar 2023 19:49:22	U	SMPL
CRA - 1	CRA	909	0.2261	-	7471B	PPB	29 Mar 2023 19:57:28	U	SMPL
HighStd - 1	HighStd	39127	11.4889	-	7471B	PPB	29 Mar 2023 19:59:43	U	SMPL
ChkStd - 1	ChkStd	27205	7.9755	-	7471B	PPB	29 Mar 2023 20:01:59	U	SMPL
PB151772BL - 1	PB151772BL	13	-0.0380	-	7471B	PPB	29 Mar 2023 20:04:18	U	SMPL
PB151772BS - 1	PB151772BS	14559	4.2487	-	7471B	PPB	29 Mar 2023 20:09:16	U	SMPL
O2105-01 - 1	WV-SOIL1-0327	7181	2.0744	-	7471B	PPB	29 Mar 2023 20:11:32	U	SMPL
O2105-04 - 1	WV-SOIL2-0327	6373	1.8363	-	7471B	PPB	29 Mar 2023 20:13:50	U	SMPL
O2105-07 - 1	WV-SOIL3-0327	1109	0.2850	-	7471B	PPB	29 Mar 2023 20:16:08	U	SMPL
O2106-01 - 1	VNJ-238	916	0.2281	-	7471B	PPB	29 Mar 2023 20:18:27	U	SMPL
O2108-02 - 1	VNJ-235	463	0.0946	-	7471B	PPB	29 Mar 2023 20:20:43	U	SMPL
CCV20 - 1	CCV20	18449	5.3951	-	7471B	PPB	29 Mar 2023 20:36:02	U	SMPL
CCB20 - 1	CCB20	134	-0.0023	-	7471B	PPB	29 Mar 2023 20:38:17	U	SMPL
O2108-02DUP - 1	VNJ-235DUP	3804	1.0792	-	7471B	PPB	29 Mar 2023 20:40:34	U	SMPL
O2108-02 - 1	VNJ-235	3667	1.0389	-	7471B	PPB	29 Mar 2023 20:46:35	U	SMPL
O2109-01 - 1	B-30-SB01	2548	0.7091	-	7471B	PPB	29 Mar 2023 20:48:51	U	SMPL
O2109-02 - 1	B-30-SB02	920	0.2293	-	7471B	PPB	29 Mar 2023 20:51:06	U	SMPL
O2113-01 - 1	72-11935	691	0.1618	-	7471B	PPB	29 Mar 2023 20:53:22	U	SMPL
O2113-03 - 1	VNJ-206	600	0.1350	-	7471B	PPB	29 Mar 2023 20:55:39	U	SMPL
O2113-05 - 1	ETGI-287	1095	0.2809	-	7471B	PPB	29 Mar 2023 20:57:56	U	SMPL
O2113-07 - 1	RBR-200058	466	0.0955	-	7471B	PPB	29 Mar 2023 21:00:13	U	SMPL
O2117-01 - 1	NB-8-032823	1863	0.5072	-	7471B	PPB	29 Mar 2023 21:02:30	U	SMPL
CCV21 - 1	CCV21	18393	5.3786	-	7471B	PPB	29 Mar 2023 21:04:46	U	SMPL
CCB21 - 1	CCB21	173	0.0092	-	7471B	PPB	29 Mar 2023 21:07:01	U	SMPL
O2108-02MS - 1	VNJ-235MS	16073	4.6949	-	7471B	PPB	29 Mar 2023 21:09:18	U	SMPL
O2108-02MSD - 1	VNJ-235MSD	17451	5.1010	-	7471B	PPB	29 Mar 2023 21:11:34	U	SMPL
O2108-02LX5 - 1		750	0.1792	-	7471B	PPB	29 Mar 2023 21:13:50	U	SMPL
O2108-02A - 1		17182	5.0217	-	7471B	PPB	29 Mar 2023 21:16:06	U	SMPL
CCV22 - 1	CCV22	17078	4.9911	-	7471B	PPB	29 Mar 2023 21:18:22	U	SMPL
CCB22 - 1	CCB22	131	-0.0032	-	7471B	PPB	29 Mar 2023 21:20:40	U	SMPL