



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 3/31/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:20  
In Date: 03/30/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 03/31/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB124755

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O2130-01 | C1190           | 1      | 1.13           | 7.65         | 8.78                    | 7.98                      | 89.5    |          |
| O2130-02 | C1191           | 2      | 1.16           | 8.81         | 9.97                    | 8.7                       | 85.6    |          |
| O2130-03 | C1192           | 3      | 1.18           | 8.79         | 9.97                    | 9.11                      | 90.2    |          |
| O2130-04 | C1193           | 4      | 1.15           | 8.40         | 9.55                    | 8.33                      | 85.5    |          |
| O2130-05 | C1194           | 5      | 1.13           | 8.76         | 9.89                    | 8.62                      | 85.5    |          |
| O2130-06 | C1195           | 6      | 1.13           | 8.68         | 9.81                    | 8.52                      | 85.1    |          |
| O2130-07 | C1196           | 7      | 1.15           | 8.62         | 9.77                    | 8.44                      | 84.6    |          |
| O2130-08 | C1197           | 8      | 1.16           | 8.82         | 9.98                    | 9.47                      | 94.2    |          |
| O2130-09 | C1198           | 9      | 1.12           | 8.84         | 9.96                    | 8.94                      | 88.5    |          |
| O2130-10 | C1199           | 10     | 1.19           | 8.42         | 9.61                    | 8.44                      | 86.1    |          |
| O2130-11 | C11A0           | 11     | 1.15           | 8.82         | 9.97                    | 9.42                      | 93.8    |          |
| O2130-12 | C11A1           | 12     | 1.19           | 8.79         | 9.98                    | 9.44                      | 93.9    |          |
| O2130-13 | C11A2           | 13     | 1.15           | 8.83         | 9.98                    | 9.33                      | 92.6    |          |
| O2130-14 | C11A3           | 14     | 1.13           | 8.80         | 9.93                    | 9.07                      | 90.2    |          |
| O2130-15 | C11B5           | 15     | 1.12           | 8.50         | 9.62                    | 8.59                      | 87.9    |          |
| O2130-19 | VHBLK002        | 16     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | VHBLK    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$