


LB124763 INSTRUMENT: CU1

SFAM01.1

μ Abs.:

26040



Linear

A= 0.0000e+000

B= 3.7811e-004

C= -4.2729e-002

Rho= 0.9994345

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>1. D</i>
0.00	0.000	-0.028	-0.028	39	0.000	39					<i>1</i>
0.05	0.050					185	0				<i>-13</i>
0.20	0.200	0.174	-0.026	572	0.0 %	572					<i>-2</i>
2.50	2.500	2.460	-0.040	6618	0.0 %	6618					<i>2</i>
5.00	5.000	5.082	0.082	13554	0.0 %	13554					<i>3</i>
7.50	7.500	7.709	0.209	20502	0.0 %	20502					<i>-2</i>
10.0	10.000	9.803	-0.197	26040	0.0 %	26040					

LB124763 INSTRUMENT ID: CVI

Sample ID	Extended II μ Abs.	Conc.	Stnd Cor Method	Units	Date	Type
0	50.0	39 -	0 SFAM01.1	PPB	3/30/2023 19:30	Std
0.2	50.2	572 -	0.2 SFAM01.1	PPB	3/30/2023 19:32	Std
2.5	52.5	6618 -	2.5 SFAM01.1	PPB	3/30/2023 19:34	Std
5	55.0	13554 -	5 SFAM01.1	PPB	3/30/2023 19:36	Std
7.5	57.5	20502 -	7.5 SFAM01.1	PPB	3/30/2023 19:39	Std
10	510.0	26040 -	10 SFAM01.1	PPB	3/30/2023 19:41	Std
ICV007	ICV007	11175 4.1827 -	SFAM01.1	PPB	3/30/2023 19:44	SMPL
ICB007	ICB007	35 -0.0295 -	SFAM01.1	PPB	3/30/2023 19:46	SMPL
CCV026	CCV026	12493 4.681 -	SFAM01.1	PPB	3/30/2023 19:48	SMPL
CCB026	CCB026	26 -0.0329 -	SFAM01.1	PPB	3/30/2023 19:51	SMPL
PB151717BL	PBS717	73 -0.0151 -	SFAM01.1	PPB	3/30/2023 19:53	SMPL
O1918-01	YBLB9	693266 HIGH -	SFAM01.1	PPB	3/30/2023 19:55	
O1918-04	YBLC0	23341 8.7828 -	SFAM01.1	PPB	3/30/2023 20:01	SMPL
O1918-05	YBLC1	27407 10.3202 -	SFAM01.1	PPB	3/30/2023 20:03	SMPL
O1918-06	YBLC2	64970 24.5232 -	SFAM01.1	PPB	3/30/2023 20:06	SMPL
O2009-01	MCB946	-916 -0.3891 -	SFAM01.1	PPB	3/30/2023 20:08	SMPL
O2009-02	MCB948	1230 0.4223 -	SFAM01.1	PPB	3/30/2023 20:10	SMPL
O2009-03	MCB963	-28 -0.0533 -	SFAM01.1	PPB	3/30/2023 20:12	SMPL
O2009-04	MCB963D	7 -0.0401 -	SFAM01.1	PPB	3/30/2023 20:15	SMPL
O2009-05	MCB963S	3128 1.14 -	SFAM01.1	PPB	3/30/2023 20:17	SMPL
O1918-02	YBLB9D	679315 HIGH -	SFAM01.1	PPB	3/30/2023 20:19	
O1918-03	YBLB9S	706776 HIGH -	SFAM01.1	PPB	3/30/2023 20:21	
O1918-01DLX50	YBLB9	20172 7.5845 -	SFAM01.1	PPB	3/30/2023 20:23	SMPL
O1918-02DLX50	YBLB9D	21818 8.2069 -	SFAM01.1	PPB	3/30/2023 20:25	SMPL
O1918-03DLX50	YBLB9S	21320 8.0186 -	SFAM01.1	PPB	3/30/2023 20:28	SMPL
O1918-05DLX2	YBLC1	12464 4.6701 -	SFAM01.1	PPB	3/30/2023 20:30	SMPL
O1918-06DLX5	YBLC2	11432 4.2798 -	SFAM01.1	PPB	3/30/2023 20:32	SMPL
CCV027	CCV027	13409 5.0274 -	SFAM01.1	PPB	3/30/2023 20:35	SMPL
CCB027	CCB027	4 -0.0412 -	SFAM01.1	PPB	3/30/2023 20:37	SMPL