



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124781

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	03/31/2023	16:10
2	CAL2	1	Water	NA	NA	20.7	7.00	03/31/2023	16:11
3	CAL3	1	Water	NA	NA	20.8	10.03	03/31/2023	16:11
4	ICV	1	Water	NA	NA	20.6	6.99	03/31/2023	16:15
5	CCV1	1	Water	NA	NA	20.7	2.01	03/31/2023	16:22
6	O2129-03	1	Water	NA	NA	20.4	7.43	03/31/2023	16:25
7	O2129-03DUP	1	Water	NA	NA	20.5	7.44	03/31/2023	16:26
8	CCV2	1	Water	NA	NA	20.7	12.02	03/31/2023	16:30

WORKLIST(Hardcopy Internal Chain)

VP 124781

WorkList Name : CORROSIVITY 02129 W WorkList ID : 168674 Department : Wet-Chemistry Date : 03-31-2023 14:39:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/30/2023	Water	02129-03	Corrosivity		Cool 4 deg C	PSEG03 M12	5730	03/30/2023	9040C

Date/Time 03/31/23 16:00
 Raw Sample Received by: JTCSC
 Raw Sample Relinquished by: JTCSC

Date/Time 03/31/23 17:35
 Raw Sample Received by: JTCSC
 Raw Sample Relinquished by: JTCSC