



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 03/31/2023 17:05 TEMP1 OUT: 103 °C 04/03/2023 08:00
TEMP2 IN: 104 °C 04/03/2023 08:30 TEMP2 OUT: 104 °C 04/03/2023 10:00
TEMP3 IN: 540 °C 04/03/2023 10:30 TEMP3 OUT: 104 °C 04/03/2023 12:00
TEMP4 IN: 560 °C 04/03/2023 12:30 TEMP4 OUT: 560 °C 04/03/2023 14:00

Run Number: LB124784

SUPERVISOR: Iwona

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN#1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB124784BL	75.9681	75.9681	100	75.9681	75.9681	75.9681	75.9681	0.0000	0
2	O2167-01	100.7404	100.7404	100	100.8139	100.8139	100.7792	100.7792	0.0347	347
3	O2167-01DUP	68.7843	68.7843	100	68.8569	68.8569	68.8208	68.8208	0.0361	361

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105 °C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

VB121284

WorkList Name : TVS O2167

WorkList ID : 168678

Department : Wet-Chemistry

Date : 03-31-2023 15:42:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/07/2023	Water	O2167-01	TVS	Cool 4 deg C	PSEG03	TW-1		03/31/2023	SM2540 G

16100

Date/Time

03/31/23

Raw Sample Received by:

70(WC)

Raw Sample Relinquished by:

ITCsm

Date/Time

03/31/23

Raw Sample Received by:

ITCsm

Raw Sample Relinquished by:

70(WC)