



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124785

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	03/31/2023	16:10
2	CAL2	1	Water	NA	NA	20.7	7.00	03/31/2023	16:11
3	CAL3	1	Water	NA	NA	20.8	10.03	03/31/2023	16:11
4	ICV	1	Water	NA	NA	20.6	6.99	03/31/2023	16:15
5	CCV1	1	Water	NA	NA	20.7	2.01	03/31/2023	16:22
6	O2167-01	1	Water	NA	NA	20.6	6.80	03/31/2023	17:10
7	O2167-01DUP	1	Water	NA	NA	20.6	6.81	03/31/2023	17:11
8	CCV2	1	Water	NA	NA	20.7	12.02	03/31/2023	17:15

VB 124785

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 02167		WorkList ID : 168681		Department : Wet-Chemistry		Date : 03-31-2023 15:58:14			
Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/31/2023	Water	02167-01	pH		Cool 4 deg C	PSEG03	TW-1	03/31/2023	9040C

Date/Time 03.31.23 16:10
Raw Sample Received by: JDC (sm)
Raw Sample Relinquished by: JDC (sm)

Date/Time 03.31.23 17:50
Raw Sample Received by: JDC (sm)
Raw Sample Relinquished by: JDC (sm)