


LB124788, INSTRUMENT 1 CV1

7471B

Linear

μ Abs.:

58497



A= 0.0000e+000

B= 1.7232e-004

C= -6.8947e-002

Rho= 0.9999890

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	378	0.000	378				3
0.2	0.200	0.205	0.005	1587	0.0 %	1587				1
2.5	2.500	2.522	0.022	15036	0.0 %	15036				-1
5.0	5.000	4.966	-0.034	29219	0.0 %	29219				0
7.5	7.500	7.500	-0.000	43922	0.0 %	43922				0
10.0	10.000	10.011	0.011	58497	0.0 %	58497				

LB124788, INSTRUMENT ID:CV1

Method: #7471B		Operator: Admin		Date of Analysis: 02 Apr 2023 13:39:21					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	378	-	0.0000	7471B	PPB	02 Apr 2023 13:40:50	S	Std
0.2 - 1	50.2	1587	-	0.2000	7471B	PPB	02 Apr 2023 13:43:07	S	Std
2.5 - 1	52.5	15036	-	2.5000	7471B	PPB	02 Apr 2023 13:45:25	S	Std
5.0 - 1	55.0	29219	-	5.0000	7471B	PPB	02 Apr 2023 13:47:42	S	Std
7.5 - 1	57.5	43922	-	7.5000	7471B	PPB	02 Apr 2023 13:49:58	S	Std
10.0 - 1	510.0	58497	-	10.0000	7471B	PPB	02 Apr 2023 13:52:16	S	Std
ICV114 - 1	ICV114	25297	4.2903	-	7471B	PPB	02 Apr 2023 13:55:09	U	SMPL
ICB114 - 1	ICB114	284	-0.0200	-	7471B	PPB	02 Apr 2023 13:57:24	U	SMPL
CCV258 - 1	CCV258	29198	4.9625	-	7471B	PPB	02 Apr 2023 13:59:41	U	SMPL
CCB258 - 1	CCB258	283	-0.0202	-	7471B	PPB	02 Apr 2023 14:01:56	U	SMPL
CRA - 1	CRA	1557	0.1994	-	7471B	PPB	02 Apr 2023 14:04:13	U	SMPL
HighStd - 1	HighStd	58336	9.9836	-	7471B	PPB	02 Apr 2023 14:06:29	U	SMPL
ChkStd - 1	ChkStd	40069	6.8358	-	7471B	PPB	02 Apr 2023 14:08:45	U	SMPL
PB151848BL - 1	PB151848BL	213	-0.0322	-	7471B	PPB	02 Apr 2023 14:11:03	U	SMPL
PB151848BS - 1	PB151848BS	25330	4.2960	-	7471B	PPB	02 Apr 2023 14:13:20	U	SMPL
O2126-01 - 1	RS-A	1641	0.2138	-	7471B	PPB	02 Apr 2023 14:15:37	U	SMPL
O2126-02 - 1	RS-B	3345	0.5075	-	7471B	PPB	02 Apr 2023 14:17:55	U	SMPL
O2126-03 - 1	RS-C	17777	2.9944	-	7471B	PPB	02 Apr 2023 14:20:13	U	SMPL
O2126-04 - 1	RS-D	6125	0.9865	-	7471B	PPB	02 Apr 2023 14:22:30	U	SMPL
O2126-05 - 1	RS-E	5271	0.8394	-	7471B	PPB	02 Apr 2023 14:24:46	U	SMPL
CCV29 - 1	CCV29	28992	4.9270	-	7471B	PPB	02 Apr 2023 14:27:02	U	SMPL
CCB29 - 1	CCB29	288	-0.0193	-	7471B	PPB	02 Apr 2023 14:29:17	U	SMPL
O2126-06 - 1	RS-F	36785	6.2699	-	7471B	PPB	02 Apr 2023 14:31:36	U	SMPL
O2133-01 - 1	HD-02-033023	1877	0.2545	-	7471B	PPB	02 Apr 2023 14:33:51	U	SMPL
O2136-01 - 1	SU-03-033023	1140	0.1275	-	7471B	PPB	02 Apr 2023 14:36:08	U	SMPL
O2136-01DUP - 1	SU-03-033023DUP	1041	0.1104	-	7471B	PPB	02 Apr 2023 14:38:24	U	SMPL
O2136-01MS - 1	SU-03-033023MS	25275	4.2865	-	7471B	PPB	02 Apr 2023 14:40:40	U	SMPL
O2136-01MSD - 1	SU-03-033023MSD	25417	4.3109	-	7471B	PPB	02 Apr 2023 14:42:57	U	SMPL
O2148-01 - 1	72-11984	566	0.0286	-	7471B	PPB	02 Apr 2023 14:45:15	U	SMPL
O2148-03 - 1	72-12017	705	0.0525	-	7471B	PPB	02 Apr 2023 14:47:33	U	SMPL
O2148-05 - 1	ETGI-288	237	-0.0281	-	7471B	PPB	02 Apr 2023 14:49:50	U	SMPL
O2154-01 - 1	CONCRETE-SECTION-CURB	368	-0.0055	-	7471B	PPB	02 Apr 2023 14:52:06	U	SMPL
CCV30 - 1	CCV30	29174	4.9584	-	7471B	PPB	02 Apr 2023 14:54:22	U	SMPL
CCB30 - 1	CCB30	302	-0.0169	-	7471B	PPB	02 Apr 2023 14:56:38	U	SMPL
O2161-01 - 1	WC-6	3509	0.5357	-	7471B	PPB	02 Apr 2023 14:58:55	U	SMPL
O2161-04 - 1	WC-5	2018	0.2788	-	7471B	PPB	02 Apr 2023 15:01:11	U	SMPL
O2162-01 - 1	OR-3-033123	996	0.1027	-	7471B	PPB	02 Apr 2023 15:03:27	U	SMPL
O2166-01 - 1	HR-01-033123	2335	0.3334	-	7471B	PPB	02 Apr 2023 15:05:43	U	SMPL
O2166-03 - 1	HR-02-033123	3944	0.6107	-	7471B	PPB	02 Apr 2023 15:07:59	U	SMPL
O2166-05 - 1	HR-03-033123	3003	0.4485	-	7471B	PPB	02 Apr 2023 15:10:15	U	SMPL
O2136-01A - 1		25276	4.2866	-	7471B	PPB	02 Apr 2023 15:14:49	U	SMPL
CCV31 - 1	CCV31	29119	4.9489	-	7471B	PPB	02 Apr 2023 15:17:06	U	SMPL
CCB31 - 1	CCB31	297	-0.0178	-	7471B	PPB	02 Apr 2023 15:19:23	U	SMPL
O2136-01LX5 - 1		432	0.0055	-	7471B	PPB	02 Apr 2023 15:24:03	U	SMPL
CCV32 - 1	CCV32	29179	4.9592	-	7471B	PPB	02 Apr 2023 15:26:19	U	SMPL
CCB32 - 1	CCB32	293	-0.0185	-	7471B	PPB	02 Apr 2023 15:28:34	U	SMPL