


7470A

Linear

 μ Abs.:

49945



A= 0.0000e+000

B= 2.0356e-004

C= -5.9641e-002

Rho= 0.9998143

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y.D
0.0	0.000	0.026	0.026	421	0.000	421	0				✓
0.2	0.200	0.215	0.015	1350	0.0 %	1350					8
2.5	2.500	2.523	0.023	12685	0.0 %	12685					1
5.0	5.000	4.952	-0.048	24619	0.0 %	24619					1
7.5	7.500	7.377	-0.123	36533	0.0 %	36533					1
10.0	10.000	10.107	0.107	49945	0.0 %	49945					2

LB124789, INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Apr 2023 16:26:33

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50.0	421	-	0.0000	7470A	PPB	02 Apr 2023 16:31:36	S	Std
0.2 - 1	50.2	1350	-	0.2000	7470A	PPB	02 Apr 2023 16:33:53	S	Std
2.5 - 1	52.5	12685	-	2.5000	7470A	PPB	02 Apr 2023 16:36:10	S	Std
5.0 - 1	55.0	24619	-	5.0000	7470A	PPB	02 Apr 2023 16:38:28	S	Std
7.5 - 1	57.5	36533	-	7.5000	7470A	PPB	02 Apr 2023 16:40:44	S	Std
10.0 - 1	57.5	49945	-	10.0000	7470A	PPB	02 Apr 2023 16:43:00	S	Std
ICV09 - 1	ICV09	21382	4.2929	-	7470A	PPB	02 Apr 2023 16:51:12	U	SMPL
ICB09 - 1	ICB09	336	0.0088	-	7470A	PPB	02 Apr 2023 16:53:27	U	SMPL
CCV33 - 1	CCV33	24979	5.0251	-	7470A	PPB	02 Apr 2023 16:55:43	U	SMPL
CCB33 - 1	CCB33	310	0.0035	-	7470A	PPB	02 Apr 2023 16:57:58	U	SMPL
CRA - 1	CRA	1343	0.2137	-	7470A	PPB	02 Apr 2023 17:00:13	U	SMPL
HighStd - 1	HighStd	49470	10.0106	-	7470A	PPB	02 Apr 2023 17:02:28	U	SMPL
ChkStd - 1	ChkStd	33597	6.7794	-	7470A	PPB	02 Apr 2023 17:04:44	U	SMPL
PB151849BL - 1	PBW	222	-0.0145	-	7470A	PPB	02 Apr 2023 17:07:01	U	SMPL
PB151849BS - 1	LCSW	21432	4.3031	-	7470A	PPB	02 Apr 2023 17:11:58	U	SMPL
O2140-01 - 1	MW-01	7579	1.4832	-	7470A	PPB	02 Apr 2023 17:14:15	U	SMPL
O2140-01DUP - 1	MW-01DUP	7347	1.4359	-	7470A	PPB	02 Apr 2023 17:16:31	U	SMPL
O2140-01MS - 1	MW-01MS	26538	5.3425	-	7470A	PPB	02 Apr 2023 17:18:48	U	SMPL
O2140-01MSD - 1	MW-01MSD	26607	5.3565	-	7470A	PPB	02 Apr 2023 17:21:05	U	SMPL
O2140-02 - 1	MW-01	94	-0.0405	-	7470A	PPB	02 Apr 2023 17:23:22	U	SMPL
CCV34 - 1	CCV34	24893	5.0076	-	7470A	PPB	02 Apr 2023 17:25:38	U	SMPL
CCB34 - 1	CCB34	238	-0.0112	-	7470A	PPB	02 Apr 2023 17:27:53	U	SMPL
O2140-04 - 1	MW-02	4971	0.9523	-	7470A	PPB	02 Apr 2023 17:30:09	U	SMPL
O2140-05 - 1	MW-02	70	-0.0454	-	7470A	PPB	02 Apr 2023 17:32:25	U	SMPL
O2140-07 - 1	MW-02DUP	4670	0.8910	-	7470A	PPB	02 Apr 2023 17:34:41	U	SMPL
O2140-08 - 1	MW-02DUP	213	-0.0163	-	7470A	PPB	02 Apr 2023 17:36:56	U	SMPL
O2167-01 - 1	TW-1	1620	0.2701	-	7470A	PPB	02 Apr 2023 17:39:12	U	SMPL
PB151851BL - 1	PBW	358	0.0132	-	7470A	PPB	02 Apr 2023 17:41:29	U	SMPL
PB151851BS - 1	LCSW	21010	4.2172	-	7470A	PPB	02 Apr 2023 17:43:46	U	SMPL
O2113-02 - 1	72-11935	286	-0.0014	-	7470A	PPB	02 Apr 2023 17:46:03	U	SMPL
O2113-04 - 1	VNJ-206	370	0.0157	-	7470A	PPB	02 Apr 2023 17:48:20	U	SMPL
O2113-06 - 1	ETGI-287	393	0.0204	-	7470A	PPB	02 Apr 2023 17:50:36	U	SMPL
CCV35 - 1	CCV35	24982	5.0258	-	7470A	PPB	02 Apr 2023 17:52:51	U	SMPL
CCB35 - 1	CCB35	271	-0.0045	-	7470A	PPB	02 Apr 2023 17:55:07	U	SMPL
O2113-08 - 1	RBR-200058	363	0.0143	-	7470A	PPB	02 Apr 2023 17:57:23	U	SMPL
O2129-01 - 1	COMP1-0330	362	0.0140	-	7470A	PPB	02 Apr 2023 17:59:41	U	SMPL
O2129-01DUP - 1	COMP1-0330DUP	374	0.0165	-	7470A	PPB	02 Apr 2023 18:01:56	U	SMPL
O2129-01MS - 1	COMP1-0330MS	19590	3.9281	-	7470A	PPB	02 Apr 2023 18:04:12	U	SMPL
O2129-01MSD - 1	COMP1-0330MSD	19770	3.9648	-	7470A	PPB	02 Apr 2023 18:06:28	U	SMPL
O2129-02 - 1	5710	1069	0.1580	-	7470A	PPB	02 Apr 2023 18:08:45	U	SMPL
O2129-03 - 1	5730	274	-0.0039	-	7470A	PPB	02 Apr 2023 18:11:02	U	SMPL
O2129-03DUP - 1	5730DUP	309	0.0033	-	7470A	PPB	02 Apr 2023 18:13:19	U	SMPL
O2129-03MS - 1	5730MS	21811	4.3803	-	7470A	PPB	02 Apr 2023 18:25:52	U	SMPL
O2129-03MSD - 1	5730MSD	21982	4.4151	-	7470A	PPB	02 Apr 2023 18:28:07	U	SMPL
CCV36 - 1	CCV36	24796	4.9879	-	7470A	PPB	02 Apr 2023 18:30:24	U	SMPL
CCB36 - 1	CCB36	301	0.0016	-	7470A	PPB	02 Apr 2023 18:32:40	U	SMPL
O2140-01LX5 - 1		1781	0.3029	-	7470A	PPB	02 Apr 2023 18:34:56	U	SMPL
O2140-01A - 1		26546	5.3441	-	7470A	PPB	02 Apr 2023 18:37:12	U	SMPL
O2129-01LX5 - 1		306	0.0026	-	7470A	PPB	02 Apr 2023 18:39:27	U	SMPL
O2129-01A - 1		23639	4.7524	-	7470A	PPB	02 Apr 2023 18:41:43	U	SMPL
O2129-03LX5 - 1		265	-0.0057	-	7470A	PPB	02 Apr 2023 18:43:59	U	SMPL
O2129-03A - 1		23733	4.7715	-	7470A	PPB	02 Apr 2023 18:46:16	U	SMPL
O2004-01 - 1		272	-0.0043	-	7470A	PPB	02 Apr 2023 18:48:32	U	SMPL
PB151797TB - 1		318	0.0051	-	7470A	PPB	02 Apr 2023 18:50:50	U	SMPL
PB151807TB - 1		315	0.0045	-	7470A	PPB	02 Apr 2023 18:53:08	U	SMPL
CCV37 - 1	CCV37	24630	4.9541	-	7470A	PPB	02 Apr 2023 18:55:24	U	SMPL
CCB37 - 1	CCB37	300	0.0014	-	7470A	PPB	02 Apr 2023 18:57:40	U	SMPL