


LB124790, INSTRUMENT: CW,

SFAM01.1

μ Abs.:

51180



Linear

A= 0.0000e+000

B= 1.9683e-004

C= -9.4817e-002

Rho= 0.9999799

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.024	0.024	605	0.000	605				
0.05	0.050					-46	0			
0.20	0.200	0.174	-0.026	1365	0.0 %	1365				
2.50	2.500	2.499	-0.001	13176	0.0 %	13176				
5.00	5.000	4.987	-0.013	25816	0.0 %	25816				
7.50	7.500	7.537	0.037	38775	0.0 %	38775				
10.0	10.000	9.979	-0.021	51180	0.0 %	51180				

1.0
=
-13
0
0
0
0

LB124790, Instrument ID: CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	
0 50.0	605	-	0	SFAM01.1	PPB	4/2/2023 20:47	Std	
0.2 50.2	1365	-	0.2	SFAM01.1	PPB	4/2/2023 20:49	Std	
2.5 52.5	13176	-	2.5	SFAM01.1	PPB	4/2/2023 20:51	Std	
5 55.0	25816	-	5	SFAM01.1	PPB	4/2/2023 20:53	Std	
7.5 57.5	38775	-	7.5	SFAM01.1	PPB	4/2/2023 20:56	Std	
10 510.0	51180	-	10	SFAM01.1	PPB	4/2/2023 20:58	Std	
ICV010	ICV010	22459	4.3259	-	SFAM01.1	PPB	4/2/2023 21:01	SMPL
ICB010	ICB010	510	0.0056	-	SFAM01.1	PPB	4/2/2023 21:04	SMPL
CCV038	CCV038	25680	4.9599	-	SFAM01.1	PPB	4/2/2023 21:06	SMPL
CCB038	CCB038	839	0.0703	-	SFAM01.1	PPB	4/2/2023 21:08	SMPL
PB151845BL	PBW	645	0.0321	-	SFAM01.1	PPB	4/2/2023 21:10	SMPL
O2066-12	MEYF71	135	-0.0682	-	SFAM01.1	PPB	4/2/2023 21:13	SMPL
O2066-14	MEYF73	117	-0.0718	-	SFAM01.1	PPB	4/2/2023 21:15	SMPL
O2066-15	MEYF73D	169	-0.0616	-	SFAM01.1	PPB	4/2/2023 21:17	SMPL
O2066-16	MEYF73S	5337	0.9557	-	SFAM01.1	PPB	4/2/2023 21:19	SMPL
O2137-01	SY-6	185	-0.0584	-	SFAM01.1	PPB	4/2/2023 21:22	SMPL
O2137-02	SY-6	160	-0.0633	-	SFAM01.1	PPB	4/2/2023 21:24	SMPL
O2137-03	SY-3DD	172	-0.061	-	SFAM01.1	PPB	4/2/2023 21:26	SMPL
O2137-04	SY-3DD	171	-0.0612	-	SFAM01.1	PPB	4/2/2023 21:28	SMPL
O2137-05	SY-3D	243	-0.047	-	SFAM01.1	PPB	4/2/2023 21:31	SMPL
O2137-06	SY-3D	194	-0.0566	-	SFAM01.1	PPB	4/2/2023 21:33	SMPL
O2137-07	SY-2R	110	-0.0732	-	SFAM01.1	PPB	4/2/2023 21:35	SMPL
O2137-08	SY-2R	144	-0.0665	-	SFAM01.1	PPB	4/2/2023 21:38	SMPL
O2137-09	SY-2D	116	-0.072	-	SFAM01.1	PPB	4/2/2023 21:40	SMPL
O2137-10	SY-2D	116	-0.072	-	SFAM01.1	PPB	4/2/2023 21:42	SMPL
O2137-11	SY-5	128	-0.0696	-	SFAM01.1	PPB	4/2/2023 21:44	SMPL
O2137-12	SY-5	109	-0.0734	-	SFAM01.1	PPB	4/2/2023 21:47	SMPL
O2137-14	FIELD-BLAN	95	-0.0761	-	SFAM01.1	PPB	4/2/2023 21:49	SMPL
O2137-15	FIELD-BLAN	129	-0.0694	-	SFAM01.1	PPB	4/2/2023 21:51	SMPL
PB151846BL	PBW	771	0.0569	-	SFAM01.1	PPB	4/2/2023 21:53	SMPL
CCV039	CCV039	25926	5.0083	-	SFAM01.1	PPB	4/2/2023 21:56	SMPL
CCB039	CCB039	566	0.0166	-	SFAM01.1	PPB	4/2/2023 21:58	SMPL
O2094-01	MCB964	209	-0.0537	-	SFAM01.1	PPB	4/2/2023 22:00	SMPL
O2094-02	MCB964D	172	-0.061	-	SFAM01.1	PPB	4/2/2023 22:02	SMPL
O2094-03	MCB964S	5441	0.9762	-	SFAM01.1	PPB	4/2/2023 22:05	SMPL
O2094-04	MCB971	258	-0.044	-	SFAM01.1	PPB	4/2/2023 22:07	SMPL
O2094-05	MCB973	204	-0.0547	-	SFAM01.1	PPB	4/2/2023 22:09	SMPL
O2094-06	MCB975	209	-0.0537	-	SFAM01.1	PPB	4/2/2023 22:12	SMPL
O2094-07	MCB978	296	-0.0366	-	SFAM01.1	PPB	4/2/2023 22:14	SMPL
O2094-08	MCB981	97	-0.0757	-	SFAM01.1	PPB	4/2/2023 22:16	SMPL
O2094-09	MCB982	312	-0.0334	-	SFAM01.1	PPB	4/2/2023 22:18	SMPL
O2094-10	MCB984	260	-0.0436	-	SFAM01.1	PPB	4/2/2023 22:21	SMPL
O2094-11	MCB986	327	-0.0305	-	SFAM01.1	PPB	4/2/2023 22:23	SMPL
O2094-14	MCB989	342	-0.0275	-	SFAM01.1	PPB	4/2/2023 22:25	SMPL
O2094-15	MCB991	391	-0.0179	-	SFAM01.1	PPB	4/2/2023 22:27	SMPL
O2094-16	MCB994	347	-0.0265	-	SFAM01.1	PPB	4/2/2023 22:30	SMPL

O2094-17	MCB996	301	-0.0356 -	SFAM01.1 PPB	4/2/2023 22:32 SMPL
O2094-18	MCB998	318	-0.0322 -	SFAM01.1 PPB	4/2/2023 22:34 SMPL
O2094-19	MCB9A0	270	-0.0417 -	SFAM01.1 PPB	4/2/2023 22:37 SMPL
O2094-20	MCB9A2	282	-0.0393 -	SFAM01.1 PPB	4/2/2023 22:39 SMPL
O2094-21	MCB9A4	321	-0.0316 -	SFAM01.1 PPB	4/2/2023 22:41 SMPL
O2094-22	MCB9A6	267	-0.0423 -	SFAM01.1 PPB	4/2/2023 22:43 SMPL
CCV040	CCV040	25615	4.9471 -	SFAM01.1 PPB	4/2/2023 22:46 SMPL
CCB040	CCB040	556	0.0146 -	SFAM01.1 PPB	4/2/2023 22:48 SMPL
PB151847BL	PBW	615	0.0262 -	SFAM01.1 PPB	4/2/2023 22:50 SMPL
O2095-01	MCB965	467	-0.0029 -	SFAM01.1 PPB	4/2/2023 22:52 SMPL
O2095-02	MCB965D	455	-0.0053 -	SFAM01.1 PPB	4/2/2023 22:55 SMPL
O2095-03	MCB965S	5700	1.0271 -	SFAM01.1 PPB	4/2/2023 22:57 SMPL
O2095-04	MCB968	476	-0.0011 -	SFAM01.1 PPB	4/2/2023 22:59 SMPL
O2095-05	MCB969	387	-0.0186 -	SFAM01.1 PPB	4/2/2023 23:01 SMPL
O2095-06	MCB972	421	-0.0119 -	SFAM01.1 PPB	4/2/2023 23:04 SMPL
O2095-07	MCB974	412	-0.0137 -	SFAM01.1 PPB	4/2/2023 23:06 SMPL
O2095-08	MCB976	384	-0.0192 -	SFAM01.1 PPB	4/2/2023 23:08 SMPL
O2095-09	MCB979	418	-0.0125 -	SFAM01.1 PPB	4/2/2023 23:11 SMPL
O2095-10	MCB983	357	-0.0245 -	SFAM01.1 PPB	4/2/2023 23:13 SMPL
O2095-11	MCB985	307	-0.0344 -	SFAM01.1 PPB	4/2/2023 23:15 SMPL
O2095-12	MCB987	333	-0.0293 -	SFAM01.1 PPB	4/2/2023 23:17 SMPL
O2095-13	MCB990	291	-0.0375 -	SFAM01.1 PPB	4/2/2023 23:20 SMPL
O2095-14	MCB992	696	0.0422 -	SFAM01.1 PPB	4/2/2023 23:22 SMPL
O2095-15	MCB995	1261	0.1534 -	SFAM01.1 PPB	4/2/2023 23:24 SMPL
O2095-16	MCB997	1400	0.1808 -	SFAM01.1 PPB	4/2/2023 23:26 SMPL
O2095-17	MCB999	1192	0.1398 -	SFAM01.1 PPB	4/2/2023 23:29 SMPL
O2095-18	MCB9A1	1337	0.1683 -	SFAM01.1 PPB	4/2/2023 23:31 SMPL
O2095-19	MCB9A3	981	0.0983 -	SFAM01.1 PPB	4/2/2023 23:33 SMPL
CCV041	CCV041	25789	4.9813 -	SFAM01.1 PPB	4/2/2023 23:36 SMPL
CCB041	CCB041	580	0.0193 -	SFAM01.1 PPB	4/2/2023 23:38 SMPL
O2095-20	MCB9A5	1534	0.2071 -	SFAM01.1 PPB	4/2/2023 23:40 SMPL
O2095-21	MCB9A7	1221	0.1455 -	SFAM01.1 PPB	4/2/2023 23:42 SMPL
CCV042	CCV042	25776	4.9788 -	SFAM01.1 PPB	4/2/2023 23:45 SMPL
CCB042	CCB042	598	0.0229 -	SFAM01.1 PPB	4/2/2023 23:47 SMPL
CCV043	CCV043	25719	4.9676 -	SFAM01.1 PPB	4/3/2023 0:52 SMPL
CCB043	CCB043	580	0.0193 -	SFAM01.1 PPB	4/3/2023 0:54 SMPL
PB151715BL	PBW715	623	0.0278 -	SFAM01.1 PPB	4/3/2023 0:57 SMPL
O1957-01	EYG57	185	-0.0584 -	SFAM01.1 PPB	4/3/2023 1:31 SMPL
O1957-02	EYG58	123	-0.0706 -	SFAM01.1 PPB	4/3/2023 1:33 SMPL
O1957-03	EYG60	128	-0.0696 -	SFAM01.1 PPB	4/3/2023 1:35 SMPL
O1957-04	EYG61	129	-0.0694 -	SFAM01.1 PPB	4/3/2023 1:38 SMPL
CCV044	CCV044	26251	5.0723 -	SFAM01.1 PPB	4/3/2023 1:40 SMPL
CCB044	CCB044	588	0.0209 -	SFAM01.1 PPB	4/3/2023 1:42 SMPL
O1957-05	EYG62	91	-0.0769 -	SFAM01.1 PPB	4/3/2023 1:44 SMPL
O1957-06	EYG63	128	-0.0696 -	SFAM01.1 PPB	4/3/2023 1:47 SMPL
O1957-07	EYG69	109	-0.0734 -	SFAM01.1 PPB	4/3/2023 1:49 SMPL
O1957-08	EYG70	142	-0.0669 -	SFAM01.1 PPB	4/3/2023 1:51 SMPL

O1957-09	EYG71	118	-0.0716 -	SFAM01.1 PPB	4/3/2023 1:54 SMPL
O1957-10	EYG77	101	-0.0749 -	SFAM01.1 PPB	4/3/2023 1:56 SMPL
O1957-11	EYG59	120	-0.0712 -	SFAM01.1 PPB	4/3/2023 1:58 SMPL
O1957-12	EYG64	113	-0.0726 -	SFAM01.1 PPB	4/3/2023 2:00 SMPL
O1957-13	EYG65	104	-0.0743 -	SFAM01.1 PPB	4/3/2023 2:03 SMPL
O1957-14	EYG66	106	-0.074 -	SFAM01.1 PPB	4/3/2023 2:05 SMPL
O1957-15	EYG67	100	-0.0751 -	SFAM01.1 PPB	4/3/2023 2:07 SMPL
O1957-16	EYG68	93	-0.0765 -	SFAM01.1 PPB	4/3/2023 2:10 SMPL
O1957-17	EYG72	115	-0.0722 -	SFAM01.1 PPB	4/3/2023 2:12 SMPL
O1957-18	EYG73	89	-0.0773 -	SFAM01.1 PPB	4/3/2023 2:14 SMPL
O1957-19	EYG73D	109	-0.0734 -	SFAM01.1 PPB	4/3/2023 2:16 SMPL
O1957-20	EYG73S	5248	0.9382 -	SFAM01.1 PPB	4/3/2023 2:19 SMPL
CCV045	CCV045	26412	5.104 -	SFAM01.1 PPB	4/3/2023 2:21 SMPL
CCB045	CCB045	582	0.0197 -	SFAM01.1 PPB	4/3/2023 2:23 SMPL