



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB124795
Analysis Date: 04/03/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: JIGNESH
REVIEWED BY: Iwona
Extraction Date: 04/03/2023
Extraction IN Time: 10:40
Extraction OUT Time: 11:15
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB124795BL	LB124795BL	WATER	1.3	1000	100	3.0259	3.0259	3.02	3.0260	3.0260	0.0001	0.1
2	LB124795BS	LB124795BS	WATER	1.3	1000	100	3.1489	3.1489	3.01	3.1657	3.1657	0.0168	16.8
3	LB124795BSD	LB124795BSD	WATER	1.3	1000	100	2.9781	2.9781	3.03	2.9951	2.9951	0.0170	17
4	O2140-03	MW-01	WATER	1.3	1000	100	3.0383	3.0383	3.05	3.0393	3.0393	0.0010	1
5	O2140-06	MW-02	WATER	1.6	1000	100	3.0540	3.0540	3.04	3.0542	3.0542	0.0002	0.2



QC Batch# LB124795 Test: Non-Polar Material Analysis Date: 04/03/2023

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3010
pH Paper 0-14	M4909
Sodium Sulfate	EP2315
1:1 HCL	WP101334
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP101332
LCSWD	5.00 ML	WP101333
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 12:31
1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time1: 11:50
Bal Check Time: 10:45 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 13:00
Out Time1: 12:30

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 14:01
1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time2: 13:30
Bal Check Time: 14:40 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 14:35
Out Time2: 14:00

124795

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPLAR O2140 W WorkList ID : 168712 Department : Wet-Chemistry Date : 04-03-2023 10:23:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/06/2023	Water	O2140-03	Non-Polar Material	Conc H2SO4 to LIRO01	M11	MW-01		03/30/2023	1664A
04/06/2023	Water	O2140-06	Non-Polar Material	Conc H2SO4 to LIRO01	M11	MW-02		03/30/2023	1664A

Date/Time 04/03/23 10:30
Raw Sample Received by: 201600
Raw Sample Relinquished by: 201600

Date/Time 04/03/23 12:30
Raw Sample Received by: 201600
Raw Sample Relinquished by: 201600