


LB124799

SFAM01.1

INSTRUMENT ID: CYJ



Linear

A= 0.0000e+000

B= 1.8907e-004 slope

C= -3.0807e-002 y-intercept

Rho= 0.9999951

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.00	0.000	0.002	0.002	172	0.000	172					
0.05	0.050	0.052	0.002	438	0.0 %	438	0				1
0.20	0.200	0.187	-0.013	1154	0.0 %	1154					4
2.50	2.500	2.495	-0.005	13358	0.0 %	13358					-7
5.00	5.000	5.022	0.022	26722	0.0 %	26722					0
7.50	7.500	7.508	0.008	39871	0.0 %	39871					0
10.0	10.000	9.985	-0.015	52974	0.0 %	52974					0

LB124799

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	172	-		0 SFAM01.1	PPB	4/3/2023 15:06	Std
	0.05 S01	438	-	0.05	SFAM01.1	PPB	4/3/2023 15:08	Std
	0.2 S02	1154	-	0.2	SFAM01.1	PPB	4/3/2023 15:10	Std
	2.5 S03	13358	-	2.5	SFAM01.1	PPB	4/3/2023 15:13	Std
	5 S04	26722	-	5	SFAM01.1	PPB	4/3/2023 15:15	Std
	7.5 S05	39871	-	7.5	SFAM01.1	PPB	4/3/2023 15:17	Std
	10 S06	52974	-	10	SFAM01.1	PPB	4/3/2023 15:24	Std
ICV001	ICV001	22681	4.2575 -		SFAM01.1	PPB	4/3/2023 15:27	SMPL
ICB001	ICB001	55	-0.0204 -		SFAM01.1	PPB	4/3/2023 15:29	SMPL
CCV001	CCV001	26714	5.02 -		SFAM01.1	PPB	4/3/2023 15:31	SMPL
CCB001	CCB001	45	-0.0223 -		SFAM01.1	PPB	4/3/2023 15:33	SMPL
PB151850BL	PBW850	151	-0.0023 -		SFAM01.1	PPB	4/3/2023 15:36	SMPL
O1994-01	EXEM5	125	-0.0072 -		SFAM01.1	PPB	4/3/2023 15:38	SMPL
O1994-02	EXEM7	83	-0.0151 -		SFAM01.1	PPB	4/3/2023 15:40	SMPL
O1994-03	EXEM9	83	-0.0151 -		SFAM01.1	PPB	4/3/2023 15:42	SMPL
O1994-04	EXEM9D	54	-0.0206 -		SFAM01.1	PPB	4/3/2023 15:45	SMPL
O1994-05	EXEM9S	5403	0.9907 -		SFAM01.1	PPB	4/3/2023 15:47	SMPL
CCV002	CCV002	26282	4.9383 -		SFAM01.1	PPB	4/3/2023 15:49	SMPL
CCB002	CCB002	91	-0.0136 -		SFAM01.1	PPB	4/3/2023 15:51	SMPL