



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124805

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/03/2023	16:15
2	CAL2	1	Water	NA	NA	20.8	7.00	04/03/2023	16:19
3	CAL3	1	Water	NA	NA	20.7	10.02	04/03/2023	16:20
4	ICV	1	Water	NA	NA	20.6	6.98	04/03/2023	16:22
5	CCV1	1	Water	NA	NA	20.7	2.01	04/03/2023	16:25
6	O2172-01	1	Water	NA	NA	21.9	4.85	04/03/2023	16:30
7	O2175-01	1	Water	NA	NA	20.3	6.59	04/03/2023	16:33
8	O2175-02	1	Water	NA	NA	20.4	8.90	04/03/2023	16:38
9	O2175-02DUP	1	Water	NA	NA	20.5	8.91	04/03/2023	16:39
10	CCV2	1	Water	NA	NA	20.7	12.03	04/03/2023	16:40

WORKLIST(Hardcopy Internal Chain)

124805

WorkList Name : PH 02175

WorkList ID : 168730

Department : Wet-Chemistry

Date : 04-03-2023 15:45:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/31/2023	Water	02172-01	pH	A	Cool 4 deg C	STER08	K11	POOL-WATER	03/31/2023 9040C
04/03/2023	Water	02175-01	pH	A	Cool 4 deg C	PSEG03	J11	AUD-1955	04/03/2023 9040C
04/03/2023	Water	02175-02	pH	A	Cool 4 deg C	PSEG03	J11	AUD-1956	04/03/2023 9040C

Date/Time 04/03/23 16:10

Raw Sample Received by: 7P1 GDC

Raw Sample Relinquished by: JDC Sm

Date/Time 04/03/23 17:15

Raw Sample Received by: JDC Sm

Raw Sample Relinquished by: JDC Sm