



## Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124807

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2744     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 8                 | 28.00      | 82.4      | 82.4            | 04/03/2023 | 12:00     |
| 2   | O2129-03    |               | 1  | 14                | 100.00     | >212.0    | >212.0          | 04/03/2023 | 12:30     |
| 3   | O2129-03DUP |               | 1  | 14                | 100.00     | >212.0    | >212.0          | 04/03/2023 | 13:00     |
| 4   | O2140-03    |               | 1  | 12                | 100.00     | >212.0    | >212.0          | 04/03/2023 | 13:30     |
| 5   | O2140-06    |               | 1  | 13                | 100.00     | >212.0    | >212.0          | 04/03/2023 | 14:00     |
| 6   | O2167-01    |               | 1  | 14                | 100.00     | >212.0    | >212.0          | 04/03/2023 | 14:30     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

6124867

WorkList Name : fp-3-30

WorkList ID : 168618

Department : Wet-Chemistry

Date : 03-30-2023 10:21:29

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|-----------------------------|-----------------|--------------|--------|
| 04/06/2023 | Water  | O2129-03 | Flash Point | Cool 4 deg C | PSEG03   | M12                         | 5730            | 03/30/2023   | 1010B  |
| 04/06/2023 | Water  | O2140-03 | Flash Point | Cool 4 deg C | LIRO01   | M11                         | MW-01           | 03/30/2023   | 1010B  |
| 04/06/2023 | Water  | O2140-06 | Flash Point | Cool 4 deg C | LIRO01   | M11                         | MW-02           | 03/30/2023   | 1010B  |
| 04/07/2023 | Water  | O2167-01 | Flash Point | Cool 4 deg C | PSEG03   | TW-1                        |                 | 03/31/2023   | 1010B  |

Date/Time 04/03/2023 11:40  
 Raw Sample Received by: RRI CWO  
 Raw Sample Relinquished by: RRI CWO

Date/Time 04/03/2023 13:10  
 Raw Sample Received by: RRI CWO  
 Raw Sample Relinquished by: RRI CWO