

LB124813


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50880



A= 0.0000e+000

B= 1.9837e-004

C= -9.9638e-003

Rho= 0.9998918

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	✓/1)
0.0	0.000	0.009	0.009	96	0.000	96					
0.05	0.050	0.054	0.004	322	0.0 %	322	0				
0.2	0.200	0.210	0.010	1111	0.0 %	1111					
2.5	2.500	2.546	0.046	12883	0.0 %	12883					
5.0	5.000	4.931	-0.069	24907	0.0 %	24907					
7.5	7.500	7.417	-0.083	37440	0.0 %	37440					
10.0	10.000	10.083	0.083	50880	0.0 %	50880					

LB124813 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 04 Apr 2023 12:06:00

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	96	-	0.0000	245.1	PPB	04 Apr 2023 12:12:06	S	Std
0.05 - 1	50.05	322	-	0.0500	245.1	PPB	04 Apr 2023 12:14:23	S	Std
0.2 - 1	50.2	1111	-	0.2000	245.1	PPB	04 Apr 2023 12:16:41	S	Std
2.5 - 1	50.5	12883	-	2.5000	245.1	PPB	04 Apr 2023 12:18:58	S	Std
5.0 - 1	5.9	24907	-	5.0000	245.1	PPB	04 Apr 2023 12:21:14	S	Std
7.5 - 1	5.75	37440	-	7.5000	245.1	PPB	04 Apr 2023 12:23:30	S	Std
10.0 - 1	5.70	50880	-	10.0000	245.1	PPB	04 Apr 2023 12:33:34	S	Std
ICV03 - 1	ICV03	20460	4.0487	-245.1		PPB	04 Apr 2023 12:39:00	U	SMPL
ICB03 - 1	ICB03	2	-0.0096	-245.1		PPB	04 Apr 2023 12:41:15	U	SMPL
CCV06 - 1	CCV06	25592	5.0667	-245.1		PPB	04 Apr 2023 12:43:31	U	SMPL
CCB06 - 1	CCB06	2	-0.0096	-245.1		PPB	04 Apr 2023 12:45:46	U	SMPL
CRA - 1	CRA	329	0.0553	-245.1		PPB	04 Apr 2023 12:48:01	U	SMPL
HighStd - 1	HighStd	51531	10.2122	-245.1		PPB	04 Apr 2023 12:50:17	U	SMPL
ChkStd - 1	ChkStd	37923	7.5128	-245.1		PPB	04 Apr 2023 12:52:32	U	SMPL
PB151892BL - 1	PBW	-85	-0.0268	-245.1		PPB	04 Apr 2023 12:54:50	U	SMPL
PB151892BS - 1	LCSW	21377	4.2306	-245.1		PPB	04 Apr 2023 12:59:41	U	SMPL
O2004-01 - 1	EFF-WASTE-WATER	1131	0.2144	-245.1		PPB	04 Apr 2023 13:06:34	U	SMPL
O2004-01DUP - 1	EFF-WASTE-WATERDUP	1066	0.2015	-245.1		PPB	04 Apr 2023 13:08:50	U	SMPL
O2004-01MS - 1	EFF-WASTE-WATERMS	19665	3.8910	-245.1		PPB	04 Apr 2023 13:11:07	U	SMPL
O2004-01MSD - 1	EFF-WASTE-WATERMSD	19300	3.8186	-245.1		PPB	04 Apr 2023 13:13:25	U	SMPL
O2011-01 - 1	EFFLUENT-DAY-1-METALS	817	0.1521	-245.1		PPB	04 Apr 2023 13:15:42	U	SMPL
CCV07 - 1	CCV07	24001	4.7511	-245.1		PPB	04 Apr 2023 13:36:20	U	SMPL
CCB07 - 1	CCB07	-52	-0.0203	-245.1		PPB	04 Apr 2023 13:38:35	U	SMPL
O2140-03 - 1	MW-01	2227	0.4318	-245.1		PPB	04 Apr 2023 13:40:52	U	SMPL
O2140-06 - 1	MW-02	5426	1.0664	-245.1		PPB	04 Apr 2023 13:43:08	U	SMPL
O2004-01LX5 - 1		238	0.0372	-245.1		PPB	04 Apr 2023 13:45:24	U	SMPL
O2004-01A - 1		20717	4.0997	-245.1		PPB	04 Apr 2023 13:47:40	U	SMPL
CCV08 - 1	CCV08	24920	4.9334	-245.1		PPB	04 Apr 2023 13:49:58	U	SMPL
CCB08 - 1	CCB08	-39	-0.0177	-245.1		PPB	04 Apr 2023 13:52:15	U	SMPL