

LB124815


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

51837



A= 0.0000e+000

B= 1.9666e-004

C= -5.3279e-002

Rho= 0.9988545

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.031	-0.031	112	0.000	112	0				
0.2	0.200	0.151	-0.049	1040	0.0 %	1040					-25
2.5	2.500	2.808	0.308	14550	0.0 %	14550					12
5.0	5.000	4.836	-0.164	24860	0.0 %	24860					-3
7.5	7.500	7.295	-0.205	37367	0.0 %	37367					-3
10.0	10.000	10.141	0.141	51837	0.0 %	51837					1

LB124815 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Apr 2023 14:47:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	112	-	0.0000	7470A	PPB	04 Apr 2023 14:49:22	S	Std
0.2 - 1	50.2	1040	-	0.2000	7470A	PPB	04 Apr 2023 14:51:39	S	Std
2.5 - 1	52.5	14550	-	2.5000	7470A	PPB	04 Apr 2023 14:53:56	S	Std
5.0 - 1	55	24860	-	5.0000	7470A	PPB	04 Apr 2023 14:56:13	S	Std
7.5 - 1	57.5	37367	-	7.5000	7470A	PPB	04 Apr 2023 14:58:30	S	Std
10.0 - 1	510	51837	-	10.0000	7470A	PPB	04 Apr 2023 15:00:47	S	Std
ICV04 - 1	ICV04	22158	4.3043	-	7470A	PPB	04 Apr 2023 15:04:25	U	SMPL
ICB04 - 1	ICB04	-158	-0.0844	-	7470A	PPB	04 Apr 2023 15:06:40	U	SMPL
CCV09 - 1	CCV09	25022	4.8675	-	7470A	PPB	04 Apr 2023 15:08:57	U	SMPL
CCB09 - 1	CCB09	-147	-0.0822	-	7470A	PPB	04 Apr 2023 15:11:13	U	SMPL
CRA - 1	CRA	1120	0.1670	-	7470A	PPB	04 Apr 2023 15:16:34	U	SMPL
HighStd - 1	HighStd	52567	10.2845	-	7470A	PPB	04 Apr 2023 15:18:49	U	SMPL
ChkStd - 1	ChkStd	34949	6.8198	-	7470A	PPB	04 Apr 2023 15:21:04	U	SMPL
PB151868BL - 1	PBW	101	-0.0334	-	7470A	PPB	04 Apr 2023 15:26:45	U	SMPL
PB151868BS - 1	LCSW	22163	4.3053	-	7470A	PPB	04 Apr 2023 15:29:01	U	SMPL
O2148-02 - 1	72-11984	-144	-0.0816	-	7470A	PPB	04 Apr 2023 15:31:18	U	SMPL
O2148-04 - 1	72-12017	33	-0.0468	-	7470A	PPB	04 Apr 2023 15:33:35	U	SMPL
O2148-06 - 1	ETGI-288	58	-0.0419	-	7470A	PPB	04 Apr 2023 15:35:53	U	SMPL
O2161-02 - 1	WC-6	51	-0.0432	-	7470A	PPB	04 Apr 2023 15:38:09	U	SMPL
O2161-05 - 1	WC-5	78	-0.0379	-	7470A	PPB	04 Apr 2023 15:40:25	U	SMPL
CCV10 - 1	CCV10	24902	4.8439	-	7470A	PPB	04 Apr 2023 15:42:41	U	SMPL
CCB10 - 1	CCB10	-122	-0.0773	-	7470A	PPB	04 Apr 2023 15:44:56	U	SMPL
O2161-05DUP - 1	WC-5DUP	72	-0.0391	-	7470A	PPB	04 Apr 2023 15:47:13	U	SMPL
O2161-05MS - 1	WC-5MS	20762	4.0298	-	7470A	PPB	04 Apr 2023 15:49:30	U	SMPL
O2161-05MSD - 1	WC-5MSD	21549	4.1845	-	7470A	PPB	04 Apr 2023 15:51:45	U	SMPL
PB151838TB - 1	PB151838TB	-137	-0.0802	-	7470A	PPB	04 Apr 2023 15:54:02	U	SMPL
O2161-05LX5 - 1		136	-0.0265	-	7470A	PPB	04 Apr 2023 15:56:20	U	SMPL
O2161-05A - 1		20818	4.0408	-	7470A	PPB	04 Apr 2023 15:58:36	U	SMPL
CCV11 - 1	CCV11	24947	4.8528	-	7470A	PPB	04 Apr 2023 16:00:53	U	SMPL
CCB11 - 1	CCB11	-160	-0.0847	-	7470A	PPB	04 Apr 2023 16:03:12	U	SMPL