



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124823

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	04/04/2023	14:40
2	O2129-02		1	14	100.00	>212.0	>212.0	04/04/2023	15:10
3	O2129-02DUP		1	14	100.00	>212.0	>212.0	04/04/2023	15:40
4	O2183-01		1	12	100.00	>212.0	>212.0	04/04/2023	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66124823

WorkList Name : FP-03-30

WorkList ID : 168621

Department : Wet-Chemistry

Date : 03-30-2023 10:22:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/06/2023	Solid	O2129-02	Flash Point	Cool 4 deg C	PSEG03	M12	5710	03/30/2023	1010B
04/11/2023	Solid	O2183-01	Flash Point	Cool 4 deg C	PSEG03	M11	34552	04/04/2023	1010B

Date/Time 04/06/2023 14:30
 Raw Sample Received by: Rng Cw12
 Raw Sample Relinquished by: JH(CDC)

Date/Time 04/04/2023 16:45
 Raw Sample Received by: JH(CDC)
 Raw Sample Relinquished by: Rng Cw12