

LB124834


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

62723



A= 0.0000e+000

B= 1.5868e-004

C= -1.3990e-002

Rho= 0.9997112

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.005	0.005	117	0.000	117					
0.2	0.200	0.175	-0.025	1192	0.0 %	1192					-13
2.5	2.500	2.415	-0.085	15305	0.0 %	15305					-5
5.0	5.000	5.187	0.187	32779	0.0 %	32779					4
7.5	7.500	7.480	-0.020	47226	0.0 %	47226					0
10.0	10.000	9.939	-0.061	62723	0.0 %	62723					-1

LB124832 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Apr 2023 10:15:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	117	-	0.0000	7471B	PPB	05 Apr 2023 10:21:19	S	Std
0.2 - 1	50.2	1192	-	0.2000	7471B	PPB	05 Apr 2023 10:23:37	S	Std
2.5 - 1	52.5	15305	-	2.5000	7471B	PPB	05 Apr 2023 10:25:54	S	Std
5.0 - 1	55	32779	-	5.0000	7471B	PPB	05 Apr 2023 10:28:11	S	Std
7.5 - 1	57.5	47226	-	7.5000	7471B	PPB	05 Apr 2023 10:30:29	S	Std
10.0 - 1	510	62723	-	10.0000	7471B	PPB	05 Apr 2023 10:35:22	S	Std
ICV05 - 1	ICV05	26085	4.1251	-	7471B	PPB	05 Apr 2023 10:38:08	U	SMPL
ICB05 - 1	ICB05	-204	-0.0464	-	7471B	PPB	05 Apr 2023 10:40:24	U	SMPL
CCV12 - 1	CCV12	33291	5.2685	-	7471B	PPB	05 Apr 2023 10:42:41	U	SMPL
CCB12 - 1	CCB12	-120	-0.0330	-	7471B	PPB	05 Apr 2023 10:44:57	U	SMPL
CRA - 1	CRA	1221	0.1798	-	7471B	PPB	05 Apr 2023 10:47:14	U	SMPL
HighStd - 1	HighStd	63880	10.1223	-	7471B	PPB	05 Apr 2023 10:49:29	U	SMPL
ChkStd - 1	ChkStd	44788	7.0928	-	7471B	PPB	05 Apr 2023 10:51:45	U	SMPL
PB151895BL - 1	PBS	-220	-0.0489	-	7471B	PPB	05 Apr 2023 10:54:03	U	SMPL
PB151895BS - 1	LCSS	24435	3.8633	-	7471B	PPB	05 Apr 2023 10:56:22	U	SMPL
O2170-01 - 1	EO-3-033123	4203	0.6529	-	7471B	PPB	05 Apr 2023 11:05:12	U	SMPL
O2170-01DUP - 1	EO-3-033123DUP	3932	0.6099	-	7471B	PPB	05 Apr 2023 11:07:30	U	SMPL
O2170-01MS - 1	EO-3-033123MS	29888	4.7286	-	7471B	PPB	05 Apr 2023 11:09:47	U	SMPL
O2170-01MSD - 1	EO-3-033123MSD	28764	4.5502	-	7471B	PPB	05 Apr 2023 11:12:05	U	SMPL
O2176-01 - 1	TR-04-040323	2471	0.3781	-	7471B	PPB	05 Apr 2023 11:14:23	U	SMPL
CCV13 - 1	CCV13	32290	5.1097	-	7471B	PPB	05 Apr 2023 11:16:41	U	SMPL
CCB13 - 1	CCB13	-71	-0.0253	-	7471B	PPB	05 Apr 2023 11:18:56	U	SMPL
O2183-01 - 1	34552	429	0.0541	-	7471B	PPB	05 Apr 2023 11:21:14	U	SMPL
O2170-01LX5 - 1		847	0.1204	-	7471B	PPB	05 Apr 2023 11:23:29	U	SMPL
O2170-01A - 1		29569	4.6779	-	7471B	PPB	05 Apr 2023 11:25:45	U	SMPL
CCV14 - 1	CCV14	32633	5.1641	-	7471B	PPB	05 Apr 2023 11:28:01	U	SMPL
CCB14 - 1	CCB14	-82	-0.0270	-	7471B	PPB	05 Apr 2023 11:30:19	U	SMPL