

LB124835


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

50857



A= 0.0000e+000

B= 1.9705e-004 *5/01/2*C= 1.9486e-003 *2/20/2009*

Rho= 0.9999831

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>9.011</i>
0.00	0.000	0.025	0.025	117	0.000	117					
0.05	0.050					438	0				
0.20	0.200	0.211	0.011	1062	0.0 %	1062					
2.50	2.500	2.469	-0.031	12521	0.0 %	12521					
5.00	5.000	4.978	-0.022	25254	0.0 %	25254					
7.50	7.500	7.493	-0.007	38018	0.0 %	38018					
10.0	10.000	10.023	0.023	50857	0.0 %	50857					

9.011
 117
 438
 1062
 12521
 25254
 38018
 50857

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	117	-		0 SFAM01.1	PPB	4/5/2023 12:02	Std
	0.2 S01	1062	-		0.2 SFAM01.1	PPB	4/5/2023 12:05	Std
	2.5 S02	12521	-		2.5 SFAM01.1	PPB	4/5/2023 12:07	Std
	5 S03	25254	-		5 SFAM01.1	PPB	4/5/2023 12:09	Std
	7.5 S04	38018	-		7.5 SFAM01.1	PPB	4/5/2023 12:11	Std
	10 S05	50857	-		10 SFAM01.1	PPB	4/5/2023 12:14	Std
ICV006	ICV006	20965	4.133 -		SFAM01.1	PPB	4/5/2023 12:17	SMPL
ICB006	ICB006	-76	-0.013 -		SFAM01.1	PPB	4/5/2023 12:19	SMPL
CCV015	CCV015	25248	4.977 -		SFAM01.1	PPB	4/5/2023 12:22	SMPL
CCB015	CCB015	-103	-0.0183 -		SFAM01.1	PPB	4/5/2023 12:24	SMPL
PB151902BL	PBW902	67	0.0152 -		SFAM01.1	PPB	4/5/2023 12:26	SMPL
O1957-01	EYG57	107	0.023 -		SFAM01.1	PPB	4/5/2023 12:28	SMPL
O1957-02	EYG58	86	0.0189 -		SFAM01.1	PPB	4/5/2023 12:31	SMPL
O1957-03	EYG60	98	0.0213 -		SFAM01.1	PPB	4/5/2023 12:33	SMPL
O1957-04	EYG61	100	0.0217 -		SFAM01.1	PPB	4/5/2023 12:35	SMPL
O1957-05	EYG62	83	0.0183 -		SFAM01.1	PPB	4/5/2023 12:37	SMPL
O1957-06	EYG63	30	0.0079 -		SFAM01.1	PPB	4/5/2023 12:40	SMPL
O1957-07	EYG69	125	0.0266 -		SFAM01.1	PPB	4/5/2023 12:42	SMPL
O1957-08	EYG70	299	0.0609 -		SFAM01.1	PPB	4/5/2023 12:44	SMPL
O1957-09	EYG71	12	0.0043 -		SFAM01.1	PPB	4/5/2023 12:46	SMPL
O1957-10	EYG77	25	0.0069 -		SFAM01.1	PPB	4/5/2023 12:49	SMPL
O1957-11	EYG59	69	0.0155 -		SFAM01.1	PPB	4/5/2023 12:51	SMPL
O1957-12	EYG64	168	0.0351 -		SFAM01.1	PPB	4/5/2023 12:53	SMPL
O1957-13	EYG65	85	0.0187 -		SFAM01.1	PPB	4/5/2023 12:56	SMPL
O1957-14	EYG66	47	0.0112 -		SFAM01.1	PPB	4/5/2023 12:58	SMPL
O1957-15	EYG67	133	0.0282 -		SFAM01.1	PPB	4/5/2023 13:00	SMPL
O1957-16	EYG68	94	0.0205 -		SFAM01.1	PPB	4/5/2023 13:02	SMPL
O1957-17	EYG72	202	0.0418 -		SFAM01.1	PPB	4/5/2023 13:05	SMPL
O1957-18	EYG73	31	0.0081 -		SFAM01.1	PPB	4/5/2023 13:07	SMPL
O1957-19	EYG73D	122	0.026 -		SFAM01.1	PPB	4/5/2023 13:09	SMPL
CCV016	CCV016	25118	4.9514 -		SFAM01.1	PPB	4/5/2023 13:11	SMPL
CCB016	CCB016	-34	-0.0048 -		SFAM01.1	PPB	4/5/2023 13:14	SMPL
O1957-20	EYG73S	4905	0.9685 -		SFAM01.1	PPB	4/5/2023 13:16	SMPL
CCV017	CCV017	25178	4.9632 -		SFAM01.1	PPB	4/5/2023 13:18	SMPL
CCB017	CCB017	-46	-0.0071 -		SFAM01.1	PPB	4/5/2023 13:21	SMPL