



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 4/6/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 04/05/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:32
Out Date: 04/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB124840

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2196-01	GBSW9	1	1.13	8.47	9.6	6.86	67.7	
O2196-02	GBSW9D	2	1.13	8.47	9.6	6.86	67.7	
O2196-03	GBSW9S	3	1.13	8.47	9.6	6.86	67.7	
O2196-04	GBSX0	4	1.13	8.52	9.65	6.85	67.1	
O2196-05	GBSX1	5	1.18	8.64	9.82	6.86	65.7	
O2196-06	GBSX2	6	1.15	8.78	9.93	7.6	73.5	
O2196-07	GBSX3	7	1.13	8.84	9.97	8.13	79.2	
O2196-08	GBSX4	8	1.12	8.42	9.54	7.98	81.5	
O2196-09	GBSX5	9	1.16	8.49	9.65	7.67	76.7	
O2196-10	GBSX6	10	1.11	8.63	9.74	5.97	56.3	
O2196-11	GBSX7	11	1.15	8.83	9.98	6.26	57.9	
O2196-12	GBSX8	12	1.17	8.51	9.68	6.76	65.7	
O2196-13	GBSX9	13	1.16	8.82	9.98	7.28	69.4	
O2196-14	GBSY0	14	1.12	8.81	9.93	6.98	66.5	
O2196-15	GBSY1	15	1.19	8.45	9.64	6.88	67.3	
O2196-16	GBSY2	16	1.15	8.82	9.97	5.87	53.5	
O2196-17	GBSY3	17	1.15	8.77	9.92	7.81	75.9	
O2196-18	GBSY4	18	1.11	8.47	9.58	7.69	77.7	
O2196-19	GBSY5	19	1.14	8.83	9.97	8.38	82.0	
O2196-20	GBSY6	20	1.13	8.40	9.53	5.27	49.3	
O2196-21	GBSY7	21	1.18	8.42	9.6	6.15	59.0	
O2196-22	GBT08	22	1.18	8.78	9.96	7.27	69.4	
O2196-23	GBT09	23	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$