



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124848

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: FLASPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	04/05/2023	14:35
2	O2175-01		1	14	100.00	>212.0	>212.0	04/05/2023	15:05
3	O2175-02		1	14	100.00	>212.0	>212.0	04/05/2023	15:35
4	O2175-02DUP		1	12	100.00	>212.0	>212.0	04/05/2023	16:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66124848

WorkList Name : fp-4-4

WorkList ID : 168737

Department : Wet-Chemistry

Date : 04-04-2023 08:19:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/10/2023	Water	O2175-01	Flash Point	Cool 4 deg C	PSEG03	J11	AUD-1955	04/03/2023	1010B
04/10/2023	Water	O2175-02	Flash Point	Cool 4 deg C	PSEG03	J11	AUD-1956	04/03/2023	1010B

Date/Time 04/05/2023 14:20
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: SB(CWC)

Date/Time 04/05/2023 16:45
 Raw Sample Received by: SB(CWC)
 Raw Sample Relinquished by: RM CWC