

LB124862


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

49854



A= 0.0000e+000

B= 2.0130e-004 slope

C= -4.2468e-002 y-intercept

Rho= 0.9999155

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.034	0.034	378	0.000	378					
0.05	0.050					438	0				
0.20	0.200	0.218	0.018	1294	0.0 %	1294					
2.50	2.500	2.460	-0.040	12434	0.0 %	12434					
5.00	5.000	4.924	-0.076	24673	0.0 %	24673					
7.50	7.500	7.571	0.071	37820	0.0 %	37820					
10.0	10.000	9.993	-0.007	49854	0.0 %	49854					

0-1-2-9-11
0-1-2-9-11

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	378	-		0 SFAM01.1	PPB	4/6/2023 11:03	Std
	0.2 S01	1294	-		0.2 SFAM01.1	PPB	4/6/2023 11:05	Std
	2.5 S02	12434	-		2.5 SFAM01.1	PPB	4/6/2023 11:07	Std
	5 S03	24673	-		5 SFAM01.1	PPB	4/6/2023 11:10	Std
	7.5 S04	37820	-		7.5 SFAM01.1	PPB	4/6/2023 11:15	Std
	10 S05	49854	-		10 SFAM01.1	PPB	4/6/2023 11:20	Std
ICV007	ICV007	21412	4.2677 -		SFAM01.1	PPB	4/6/2023 11:22	SMPL
ICB007	ICB007	114	-0.0195 -		SFAM01.1	PPB	4/6/2023 11:25	SMPL
CCV018	CCV018	25631	5.117 -		SFAM01.1	PPB	4/6/2023 11:27	SMPL
CCB018	CCB018	147	-0.0129 -		SFAM01.1	PPB	4/6/2023 11:29	SMPL
PB151933BL	PBW933	164	-0.0095 -		SFAM01.1	PPB	4/6/2023 11:31	SMPL
PB151934BL	PBW934	186	-0.005 -		SFAM01.1	PPB	4/6/2023 11:38	SMPL
O2134-01	MCB9A8	50	-0.0324 -		SFAM01.1	PPB	4/6/2023 11:41	SMPL
O2134-02	MCB9B0	76	-0.0272 -		SFAM01.1	PPB	4/6/2023 11:43	SMPL
O2134-03	MCB9B6	33	-0.0358 -		SFAM01.1	PPB	4/6/2023 11:45	SMPL
O2134-04	MCB9B8	61	-0.0302 -		SFAM01.1	PPB	4/6/2023 11:48	SMPL
O2134-05	MCB9B2	351	0.0282 -		SFAM01.1	PPB	4/6/2023 11:50	SMPL
O2134-06	MCB9B4	46	-0.0332 -		SFAM01.1	PPB	4/6/2023 11:52	SMPL
O2134-07	MCB9C3	278	0.0135 -		SFAM01.1	PPB	4/6/2023 11:54	SMPL
O2134-08	MCB9C5	-51	-0.0527 -		SFAM01.1	PPB	4/6/2023 11:57	SMPL
O2134-09	MCB9C8	43	-0.0338 -		SFAM01.1	PPB	4/6/2023 11:59	SMPL
O2134-10	MCB9C9	84	-0.0256 -		SFAM01.1	PPB	4/6/2023 12:01	SMPL
O2134-13	MCB9C5D	69	-0.0286 -		SFAM01.1	PPB	4/6/2023 12:03	SMPL
O2134-14	MCB9C5S	5090	0.9821 -		SFAM01.1	PPB	4/6/2023 12:06	SMPL
PB151935BL	PBW935	-2	-0.0429 -		SFAM01.1	PPB	4/6/2023 12:08	SMPL
O2135-01	MCB9A9	66	-0.0292 -		SFAM01.1	PPB	4/6/2023 12:10	SMPL
O2135-02	MCB9B1	89	-0.0246 -		SFAM01.1	PPB	4/6/2023 12:12	SMPL
O2135-03	MCB9B7	102	-0.0219 -		SFAM01.1	PPB	4/6/2023 12:15	SMPL
O2135-04	MCB9B9	160	-0.0103 -		SFAM01.1	PPB	4/6/2023 12:17	SMPL
CCV019	CCV019	26715	5.3352 -		SFAM01.1	PPB	4/6/2023 12:27	SMPL
CCB019	CCB019	134	-0.0155 -		SFAM01.1	PPB	4/6/2023 12:29	SMPL
O2135-05	MCB9B3	78	-0.0268 -		SFAM01.1	PPB	4/6/2023 12:31	SMPL
O2135-06	MCB9B5	74	-0.0276 -		SFAM01.1	PPB	4/6/2023 12:34	SMPL
O2135-07	MCB9C4	679	0.0942 -		SFAM01.1	PPB	4/6/2023 12:36	SMPL
O2135-08	MCB9C6	110	-0.0203 -		SFAM01.1	PPB	4/6/2023 12:38	SMPL
O2135-09	MCB9C6D	197	-0.0028 -		SFAM01.1	PPB	4/6/2023 12:41	SMPL
O2135-10	MCB9C6S	5348	1.0341 -		SFAM01.1	PPB	4/6/2023 12:43	SMPL
PB151936BL	PBW936	251	0.0081 -		SFAM01.1	PPB	4/6/2023 12:45	SMPL
O2190-01	SY-3	833	0.1252 -		SFAM01.1	PPB	4/6/2023 12:52	SMPL
O2190-01D	SY-3D	658	0.09 -		SFAM01.1	PPB	4/6/2023 12:54	SMPL
O2190-01S	SY-3S	6102	1.1859 -		SFAM01.1	PPB	4/6/2023 12:57	SMPL
O2190-02	SY-3	227	0.0032 -		SFAM01.1	PPB	4/6/2023 12:59	SMPL
O1378-02	MDL-WATER-QT1	442	0.0465 -		SFAM01.1	PPB	4/6/2023 13:01	SMPL
CCV020	CCV020	24702	4.93 -		SFAM01.1	PPB	4/6/2023 13:03	SMPL
CCB020	CCB020	248	0.0075 -		SFAM01.1	PPB	4/6/2023 13:06	SMPL