

LB124871


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49849



A= 0.0000e+000

B= 1.9638e-004

C= -3.9774e-002

Rho= 0.9985881

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.008	-0.008	161	0.000	161	0				-
0.2	0.200	0.173	-0.027	1082	0.0 %	1082					-14
2.5	2.500	2.542	0.042	13149	0.0 %	13149					2
5.0	5.000	4.861	-0.139	24956	0.0 %	24956					-3
7.5	7.500	7.882	0.382	40339	0.0 %	40339					5
10.0	10.000	9.750	-0.250	49849	0.0 %	49849					-3

LB124871 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Apr 2023 15:21:41

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	161	-	0.0000	7470A	PPB	07 Apr 2023 08:05:43	S	Std
0.2 - 1	50.2	1082	-	0.2000	7470A	PPB	07 Apr 2023 08:08:06	S	Std
2.5 - 1	52.5	13149	-	2.5000	7470A	PPB	07 Apr 2023 08:10:24	S	Std
5.0 - 1	55	24956	-	5.0000	7470A	PPB	07 Apr 2023 08:12:41	S	Std
7.5 - 1	59.5	40339	-	7.5000	7470A	PPB	07 Apr 2023 08:14:57	S	Std
10.0 - 1	511	49849	-	10.0000	7470A	PPB	07 Apr 2023 08:17:13	S	Std
ICV08 - 1	ICV08	20616	4.0089	-	7470A	PPB	07 Apr 2023 08:20:12	U	SMPL
ICB08 - 1	ICB08	79	-0.0243	-	7470A	PPB	07 Apr 2023 08:25:11	U	SMPL
CCV21 - 1	CCV21	24471	4.7659	-	7470A	PPB	07 Apr 2023 08:27:26	U	SMPL
CCB21 - 1	CCB21	74	-0.0252	-	7470A	PPB	07 Apr 2023 08:32:13	U	SMPL
CRA - 1	CRA	1055	0.1674	-	7470A	PPB	07 Apr 2023 08:42:29	U	SMPL
HighStd - 1	HighStd	47592	9.3065	-	7470A	PPB	07 Apr 2023 08:44:44	U	SMPL
ChkStd - 1	ChkStd	35562	6.9440	-	7470A	PPB	07 Apr 2023 08:47:00	U	SMPL
PB151951BL - 1	PBW	108	-0.0186	-	7470A	PPB	07 Apr 2023 08:51:54	U	SMPL
O1627-07 - 1	LOD-MDL-WATER-01-QT1-2023	1127	0.1815	-	7470A	PPB	07 Apr 2023 08:56:48	U	SMPL
O1627-07RE - 1	LOD-MDL-WATER-01-QT1-2023RE	1027	0.1619	-	7470A	PPB	07 Apr 2023 09:01:50	U	SMPL
O1627-09 - 1	MDL-WATER-03-QT1-2023	451	0.0488	-	7470A	PPB	07 Apr 2023 09:08:43	U	SMPL
O2175-01 - 1	AUD-1955	16427	3.1862	-	7470A	PPB	07 Apr 2023 09:11:00	U	SMPL
O2175-01DUP - 1	AUD-1955DUP	15022	2.9103	-	7470A	PPB	07 Apr 2023 09:13:15	U	SMPL
CCV22 - 1	CCV22	24212	4.7150	-	7470A	PPB	07 Apr 2023 09:23:47	U	SMPL
CCB22 - 1	CCB22	93	-0.0215	-	7470A	PPB	07 Apr 2023 09:28:35	U	SMPL
O2175-01MS - 1	AUD-1955MS	34065	6.6500	-	7470A	PPB	07 Apr 2023 09:33:28	U	SMPL
O2175-01MSD - 1	AUD-1955MSD	34610	6.7570	-	7470A	PPB	07 Apr 2023 09:35:44	U	SMPL
O2175-02 - 1	AUD-1956	2448	0.4410	-	7470A	PPB	07 Apr 2023 09:38:02	U	SMPL
O2175-01LX5 - 1		3409	0.6297	-	7470A	PPB	07 Apr 2023 09:40:20	U	SMPL
O2175-01A - 1		40338	7.8819	-	7470A	PPB	07 Apr 2023 09:42:37	U	SMPL
PB151951BS - 1	LCSW	20991	4.0825	-	7470A	PPB	07 Apr 2023 09:44:55	U	SMPL
CCV23 - 1	CCV23	27011	5.2647	-	7470A	PPB	07 Apr 2023 09:51:38	U	SMPL
CCB23 - 1	CCB23	132	-0.0139	-	7470A	PPB	07 Apr 2023 09:56:25	U	SMPL