



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124873

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	04/06/2023	12:20
2	O2195-03		1	14	100.00	>212.0	>212.0	04/06/2023	12:40
3	O2195-03DUP		1	14	100.00	>212.0	>212.0	04/06/2023	13:10
4	O2195-04		1	12	100.00	>212.0	>212.0	04/06/2023	13:40
5	O2195-05		1	14	100.00	>212.0	>212.0	04/06/2023	14:10
6	O2195-06		1	12	100.00	>212.0	>212.0	04/06/2023	14:40
7	O2206-05		1	13	100.00	>212.0	>212.0	04/06/2023	15:10
8	O2206-06		1	12	100.00	>212.0	>212.0	04/06/2023	15:40
9	O2206-07		1	14	100.00	>212.0	>212.0	04/06/2023	16:10
10	O2206-08		1	14	100.00	>212.0	>212.0	04/06/2023	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6124873

WorkList Name : fp-4-6

WorkList ID : 168799

Department : Wet-Chemistry

Date : 04-06-2023 08:17:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/12/2023	Water	O2195-03	Flash Point	Cool 4 deg C	PSEG03	M11	193	04/05/2023	1010B
04/12/2023	Water	O2195-04	Flash Point	Cool 4 deg C	PSEG03	M11	195	04/05/2023	1010B
04/12/2023	Water	O2195-05	Flash Point	Cool 4 deg C	PSEG03	M11	196	04/05/2023	1010B
04/12/2023	Water	O2195-06	Flash Point	Cool 4 deg C	PSEG03	M11	197	04/05/2023	1010B
04/10/2023	Water	O2206-05	Flash Point	Cool 4 deg C	PSEG03	M11	2793-2794	04/12/2023	1010B
04/10/2023	Water	O2206-06	Flash Point	Cool 4 deg C	PSEG03	M11	2797	04/12/2023	1010B
04/10/2023	Water	O2206-07	Flash Point	Cool 4 deg C	PSEG03	M11	2785	04/12/2023	1010B
04/10/2023	Water	O2206-08	Flash Point	Cool 4 deg C	PSEG03	M11	2792	04/12/2023	1010B

Date/Time 04/06/2023 12:15
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JH(CWC)

Date/Time 04/06/2023 16:45
 Raw Sample Received by: JH(CWC)
 Raw Sample Relinquished by: RM CWC