

LB124880


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54949



A= 0.0000e+000

B= 1.8405e-004

C= -1.3892e-001

Rho= 0.9999433

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% RSD
0.0	0.000	-0.055	-0.055	458	0.000	458					10
0.2	0.200	0.220	0.020	1951	0.0 %	1951					2
2.5	2.500	2.557	0.057	14646	0.0 %	14646					1
5.0	5.000	4.970	-0.030	27757	0.0 %	27757					0
7.5	7.500	7.533	0.033	41684	0.0 %	41684					0
10.0	10.000	9.975	-0.025	54949	0.0 %	54949					0

LB124880 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Apr 2023 10:38:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	458	-	0.0000	7471B	PPB	07 Apr 2023 10:41:58	S	Std
0.2 - 1	50.2	1951	-	0.2000	7471B	PPB	07 Apr 2023 10:44:16	S	Std
2.5 - 1	52.5	14646	-	2.5000	7471B	PPB	07 Apr 2023 10:46:33	S	Std
5.0 - 1	55	27757	-	5.0000	7471B	PPB	07 Apr 2023 10:48:50	S	Std
7.5 - 1	57.5	41684	-	7.5000	7471B	PPB	07 Apr 2023 10:51:09	S	Std
10.0 - 1	510	54949	-	10.0000	7471B	PPB	07 Apr 2023 11:01:29	S	Std
ICV09 - 1	ICV09	23828	4.2467	-	7471B	PPB	07 Apr 2023 11:04:11	U	SMPL
ICB09 - 1	ICB09	413	-0.0629	-	7471B	PPB	07 Apr 2023 11:09:04	U	SMPL
CCV24 - 1	CCV24	28628	5.1302	-	7471B	PPB	07 Apr 2023 11:13:58	U	SMPL
CCB24 - 1	CCB24	462	-0.0539	-	7471B	PPB	07 Apr 2023 11:19:45	U	SMPL
CRA - 1	CRA	1921	0.2146	-	7471B	PPB	07 Apr 2023 11:21:59	U	SMPL
HighStd - 1	HighStd	54787	9.9448	-	7471B	PPB	07 Apr 2023 11:24:15	U	SMPL
ChkStd - 1	ChkStd	39278	7.0904	-	7471B	PPB	07 Apr 2023 11:28:55	U	SMPL
PB151954BL - 1	PBS	483	-0.0500	-	7471B	PPB	07 Apr 2023 11:33:47	U	SMPL
PB151954BS - 1	LCSS	23721	4.2270	-	7471B	PPB	07 Apr 2023 11:38:20	U	SMPL
O2191-02 - 1	72-11982	30421	5.4602	-	7471B	PPB	07 Apr 2023 11:40:36	U	SMPL
O2206-01 - 1	72-11929	7977	1.3293	-	7471B	PPB	07 Apr 2023 11:42:57	U	SMPL
O2208-01 - 1	RO-124035	956	0.0370	-	7471B	PPB	07 Apr 2023 11:45:17	U	SMPL
O2209-01 - 1	PSC-124055	1150	0.0727	-	7471B	PPB	07 Apr 2023 11:47:36	U	SMPL
O2222-01 - 1	E-H4(ELEV-24)	1505	0.1381	-	7471B	PPB	07 Apr 2023 11:49:52	U	SMPL
CCV25 - 1	CCV25	27425	4.9088	-	7471B	PPB	07 Apr 2023 11:52:07	U	SMPL
CCB25 - 1	CCB25	360	-0.0727	-	7471B	PPB	07 Apr 2023 11:54:23	U	SMPL
O2222-02 - 1	MID-H4(ELEV-24)	16882	2.9683	-	7471B	PPB	07 Apr 2023 11:56:40	U	SMPL
O2222-02DUP - 1	MID-H4(ELEV-24)DUP	16128	2.8295	-	7471B	PPB	07 Apr 2023 11:58:56	U	SMPL
O2222-02MS - 1	MID-H4(ELEV-24)MS	34622	6.2334	-	7471B	PPB	07 Apr 2023 12:02:05	U	SMPL
O2222-02MSD - 1	MID-H4(ELEV-24)MSD	35785	6.4475	-	7471B	PPB	07 Apr 2023 12:04:21	U	SMPL
PB151955BL - 1	PBS	278	-0.0878	-	7471B	PPB	07 Apr 2023 12:06:39	U	SMPL
PB151955BS - 1	LCSS	24218	4.3185	-	7471B	PPB	07 Apr 2023 12:08:57	U	SMPL
O2224-01 - 1	OU3-FC-BACKFILL-01	926	0.0315	-	7471B	PPB	07 Apr 2023 12:16:03	U	SMPL
O2224-01DUP - 1	OU3-FC-BACKFILL-01DUP	823	0.0126	-	7471B	PPB	07 Apr 2023 12:18:19	U	SMPL
O2224-01MS - 1	OU3-FC-BACKFILL-01MS	25416	4.5390	-	7471B	PPB	07 Apr 2023 12:20:36	U	SMPL
O2224-01MSD - 1	OU3-FC-BACKFILL-01MSD	24285	4.3308	-	7471B	PPB	07 Apr 2023 12:22:53	U	SMPL
CCV26 - 1	CCV26	28742	5.1512	-	7471B	PPB	07 Apr 2023 12:31:19	U	SMPL
CCB26 - 1	CCB26	354	-0.0738	-	7471B	PPB	07 Apr 2023 12:33:34	U	SMPL
O2224-03 - 1	OU3-FC-BACKFILL-02	2950	0.4040	-	7471B	PPB	07 Apr 2023 12:35:52	U	SMPL
O2224-05 - 1	OU3-FC-BACKFILL-03	1111	0.0656	-	7471B	PPB	07 Apr 2023 12:38:07	U	SMPL
O2224-07 - 1	OU3-FC-BACKFILL-04	802	0.0087	-	7471B	PPB	07 Apr 2023 12:40:24	U	SMPL
O2224-09 - 1	OU3-FC-BACKFILL-05	816	0.0113	-	7471B	PPB	07 Apr 2023 12:42:39	U	SMPL
O2224-11 - 1	OU3-FC-BACKFILL-06	697	-0.0106	-	7471B	PPB	07 Apr 2023 12:44:55	U	SMPL
O2224-13 - 1	OU3-FC-BACKFILL-07	685	-0.0128	-	7471B	PPB	07 Apr 2023 12:47:12	U	SMPL
O2224-15 - 1	OU3-FC-BACKFILL-08	644	-0.0204	-	7471B	PPB	07 Apr 2023 12:49:28	U	SMPL
O2224-17 - 1	OU3-FC-BACKFILL-09	611	-0.0265	-	7471B	PPB	07 Apr 2023 12:51:45	U	SMPL
O2224-19 - 1	OU3-FC-BACKFILL-10	641	-0.0209	-	7471B	PPB	07 Apr 2023 12:54:03	U	SMPL
O2224-21 - 1	OU3-FC-BACKFILL-11	646	-0.0200	-	7471B	PPB	07 Apr 2023 12:56:18	U	SMPL
CCV27 - 1	CCV27	28749	5.1524	-	7471B	PPB	07 Apr 2023 13:06:13	U	SMPL
CCB27 - 1	CCB27	453	-0.0555	-	7471B	PPB	07 Apr 2023 13:08:29	U	SMPL
O2224-23 - 1	OU3-FC-BACKFILL-12	1743	0.1819	-	7471B	PPB	07 Apr 2023 13:10:46	U	SMPL
O2224-25 - 1	OU3-FC-BACKFILL-13	803	0.0089	-	7471B	PPB	07 Apr 2023 13:13:02	U	SMPL
O2224-27 - 1	OU3-FC-BACKFILL-14	757	0.0004	-	7471B	PPB	07 Apr 2023 13:15:18	U	SMPL
O2224-29 - 1	OU3-FC-BACKFILL-15	640	-0.0211	-	7471B	PPB	07 Apr 2023 13:17:34	U	SMPL
O2230-01 - 1	OK-01-040623	1763	0.1856	-	7471B	PPB	07 Apr 2023 13:19:50	U	SMPL
O2222-02LX5 - 1		3280	0.4648	-	7471B	PPB	07 Apr 2023 13:22:07	U	SMPL
O2222-02A - 1		36600	6.5975	-	7471B	PPB	07 Apr 2023 13:28:37	U	SMPL
O2224-01LX5 - 1		267	-0.0898	-	7471B	PPB	07 Apr 2023 13:30:53	U	SMPL
O2224-01A - 1		24786	4.4230	-	7471B	PPB	07 Apr 2023 13:34:54	U	SMPL
CCV28 - 1	CCV28	29838	5.3529	-	7471B	PPB	07 Apr 2023 13:37:09	U	SMPL
CCB28 - 1	CCB28	446	-0.0568	-	7471B	PPB	07 Apr 2023 13:39:26	U	SMPL