



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: JIGNESH

Run Number: LB124889

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	04/07/2023	11:20
2	O2214-02		1	13	100.00	>212.0	>212.0	04/07/2023	11:50
3	O2214-02DUP		1	14	100.00	>212.0	>212.0	04/07/2023	12:20
4	O2214-03		1	13	100.00	>212.0	>212.0	04/07/2023	12:50
5	O2217-01		1	14	100.00	>212.0	>212.0	04/07/2023	13:20
6	O2217-02		1	14	100.00	>212.0	>212.0	04/07/2023	13:50
7	O2221-01		1	12	100.00	>212.0	>212.0	04/07/2023	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6124889

WorkList Name : fp-4-7

WorkList ID : 168839

Department : Wet-Chemistry

Date : 04-07-2023 07:57:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/12/2023	Water	O2214-02	Flash Point	Cool 4 deg C	PSEG03	M11	D2975	04/05/2023	1010B
04/12/2023	Water	O2214-03	Flash Point	Cool 4 deg C	PSEG03	M11	T2952	04/05/2023	1010B
04/12/2023	Water	O2217-01	Flash Point	Cool 4 deg C	PSEG01	M11	351	04/05/2023	1010B
04/12/2023	Water	O2217-02	Flash Point	Cool 4 deg C	PSEG01	M11	353	04/05/2023	1010B
04/11/2023	Water	O2221-01	Flash Point	Cool 4 deg C	PSEG03	M11	NWB-1904	04/06/2023	1010B

Date/Time 04/07/2023 11:10
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: RM CWG

Date/Time 04/07/2023 15:00
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: RM CWG