


LB124890

7470A

INSTRUMENT ID: C41

Linear



A= 0.0000e+000

B= 1.9337e-004

C= -1.3009e-002

Rho= 0.9998458

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.015	0.015	147	0.000	147	0				1
0.2	0.200	0.204	0.004	1124	0.0 %	1124					2
2.5	2.500	2.568	0.068	13346	0.0 %	13346					3
5.0	5.000	4.891	-0.109	25359	0.0 %	25359					2
7.5	7.500	7.447	-0.053	38580	0.0 %	38580					1
10.0	10.000	10.075	0.075	52169	0.0 %	52169					1

LB124890 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Apr 2023 15:21:45

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	147	-	0.0000	7470A	PPB	07 Apr 2023 15:22:59	S	Std
0.2 - 1	50.2	1124	-	0.2000	7470A	PPB	07 Apr 2023 15:25:17	S	Std
2.5 - 1	52.5	13346	-	2.5000	7470A	PPB	07 Apr 2023 15:27:34	S	Std
5.0 - 1	55	25359	-	5.0000	7470A	PPB	07 Apr 2023 15:29:51	S	Std
7.5 - 1	57.5	38580	-	7.5000	7470A	PPB	07 Apr 2023 15:32:08	S	Std
10.0 - 1	510	52169	-	10.0000	7470A	PPB	07 Apr 2023 15:34:25	S	Std
ICV10 - 1	ICV10	21244	4.0949	-	7470A	PPB	07 Apr 2023 15:39:44	U	SMPL
ICB10 - 1	ICB10	-42	-0.0211	-	7470A	PPB	07 Apr 2023 15:42:00	U	SMPL
CCV29 - 1	CCV29	25580	4.9333	-	7470A	PPB	07 Apr 2023 15:44:16	U	SMPL
CCB29 - 1	CCB29	-47	-0.0221	-	7470A	PPB	07 Apr 2023 15:46:31	U	SMPL
CRA - 1	CRA	1044	0.1889	-	7470A	PPB	07 Apr 2023 15:48:47	U	SMPL
HighStd - 1	HighStd	51457	9.9371	-	7470A	PPB	07 Apr 2023 15:51:02	U	SMPL
ChkStd - 1	ChkStd	35298	6.8125	-	7470A	PPB	07 Apr 2023 15:53:18	U	SMPL
PB151980BL - 1	PBW	-167	-0.0453	-	7470A	PPB	07 Apr 2023 15:55:36	U	SMPL
PB151980BS - 1	LCSW	17845	3.4376	-	7470A	PPB	07 Apr 2023 15:57:55	U	SMPL
O2195-01 - 1	194	-36	-0.0200	-	7470A	PPB	07 Apr 2023 16:00:11	U	SMPL
O2195-02 - 1	199	36	-0.0060	-	7470A	PPB	07 Apr 2023 16:02:28	U	SMPL
O2206-03 - 1	2795	130	0.0121	-	7470A	PPB	07 Apr 2023 16:04:45	U	SMPL
O2206-04 - 1	2796	168	0.0195	-	7470A	PPB	07 Apr 2023 16:07:03	U	SMPL
O2208-02 - 1	RO-124035	100	0.0063	-	7470A	PPB	07 Apr 2023 16:09:18	U	SMPL
CCV30 - 1	CCV30	24851	4.7924	-	7470A	PPB	07 Apr 2023 16:11:34	U	SMPL
CCB30 - 1	CCB30	-19	-0.0167	-	7470A	PPB	07 Apr 2023 16:13:50	U	SMPL
O2209-02 - 1	PSC-124055	35	-0.0062	-	7470A	PPB	07 Apr 2023 16:16:08	U	SMPL
O2214-01 - 1	D2964	88	0.0040	-	7470A	PPB	07 Apr 2023 16:18:24	U	SMPL
O2214-01DUP - 1	D2964DUP	121	0.0104	-	7470A	PPB	07 Apr 2023 16:20:39	U	SMPL
O2214-01MS - 1	D2964MS	17238	3.3203	-	7470A	PPB	07 Apr 2023 16:22:55	U	SMPL
O2214-01MSD - 1	D2964MSD	17424	3.3562	-	7470A	PPB	07 Apr 2023 16:25:11	U	SMPL
PB151981BL - 1	PBW	-28	-0.0184	-	7470A	PPB	07 Apr 2023 16:27:28	U	SMPL
PB151981BS - 1	LCSW	16844	3.2441	-	7470A	PPB	07 Apr 2023 16:29:46	U	SMPL
O2195-03 - 1	193	-63	-0.0252	-	7470A	PPB	07 Apr 2023 16:32:03	U	SMPL
O2195-04 - 1	195	48	-0.0037	-	7470A	PPB	07 Apr 2023 16:34:21	U	SMPL
O2195-05 - 1	196	80	0.0025	-	7470A	PPB	07 Apr 2023 16:36:36	U	SMPL
CCV31 - 1	CCV31	24916	4.8050	-	7470A	PPB	07 Apr 2023 16:38:52	U	SMPL
CCB31 - 1	CCB31	-74	-0.0273	-	7470A	PPB	07 Apr 2023 16:41:07	U	SMPL
O2195-06 - 1	197	29	-0.0074	-	7470A	PPB	07 Apr 2023 16:43:26	U	SMPL
O2206-05 - 1	2793-2794	154	0.0168	-	7470A	PPB	07 Apr 2023 16:45:42	U	SMPL
O2206-06 - 1	2797	358	0.0562	-	7470A	PPB	07 Apr 2023 16:47:58	U	SMPL
O2206-07 - 1	2785	152	0.0164	-	7470A	PPB	07 Apr 2023 16:50:14	U	SMPL
O2206-08 - 1	2792	309	0.0467	-	7470A	PPB	07 Apr 2023 16:52:31	U	SMPL
O2214-02 - 1	D2975	43	-0.0047	-	7470A	PPB	07 Apr 2023 16:54:47	U	SMPL
O2214-03 - 1	T2952	78	0.0021	-	7470A	PPB	07 Apr 2023 16:57:04	U	SMPL
O2217-01 - 1	351	98	0.0059	-	7470A	PPB	07 Apr 2023 16:59:21	U	SMPL
O2217-02 - 1	353	168	0.0195	-	7470A	PPB	07 Apr 2023 17:01:38	U	SMPL
O2217-02DUP - 1	353DUP	112	0.0086	-	7470A	PPB	07 Apr 2023 17:03:54	U	SMPL
CCV32 - 1	CCV32	24727	4.7684	-	7470A	PPB	07 Apr 2023 17:08:52	U	SMPL
CCB32 - 1	CCB32	-109	-0.0341	-	7470A	PPB	07 Apr 2023 17:11:07	U	SMPL
O2217-02MS - 1	353MS	18473	3.5591	-	7470A	PPB	07 Apr 2023 17:16:00	U	SMPL
O2217-02MSD - 1	353MSD	18362	3.5376	-	7470A	PPB	07 Apr 2023 17:18:16	U	SMPL
PB151931TB - 1	PB151931TB	-92	-0.0308	-	7470A	PPB	07 Apr 2023 17:20:34	U	SMPL
PB151932TB - 1	PB151932TB	56	-0.0022	-	7470A	PPB	07 Apr 2023 17:22:50	U	SMPL
O2214-01LX5 - 1		93	0.0050	-	7470A	PPB	07 Apr 2023 17:25:06	U	SMPL
O2214-01A - 1		18873	3.6364	-	7470A	PPB	07 Apr 2023 17:27:23	U	SMPL
O2217-02LX5 - 1		84	0.0032	-	7470A	PPB	07 Apr 2023 17:29:40	U	SMPL
O2217-02A - 1		20293	3.9110	-	7470A	PPB	07 Apr 2023 17:31:58	U	SMPL
CCV33 - 1	CCV33	24850	4.7922	-	7470A	PPB	07 Apr 2023 17:34:15	U	SMPL
CCB33 - 1	CCB33	-124	-0.0370	-	7470A	PPB	07 Apr 2023 17:36:31	U	SMPL