



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124895

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	04/07/2023	16:10
2	CAL2	1	Water	NA	NA	20.7	7.00	04/07/2023	16:11
3	CAL3	1	Water	NA	NA	20.7	10.03	04/07/2023	16:12
4	ICV	1	Water	NA	NA	20.7	6.98	04/07/2023	16:15
5	CCV1	1	Water	NA	NA	20.7	2.01	04/07/2023	16:20
6	O2221-01	1	Water	NA	NA	20.3	6.86	04/07/2023	16:25
7	O2244-01	1	Water	NA	NA	20.1	3.87	04/07/2023	16:30
8	O2244-02	1	Water	NA	NA	20.4	4.07	04/07/2023	16:35
9	O2257-03	1	Water	NA	NA	20.9	8.26	04/07/2023	16:45
10	O2257-03DUP	1	Water	NA	NA	20.8	8.27	04/07/2023	16:46
11	CCV2	1	Water	NA	NA	20.7	12.03	04/07/2023	16:50

UB 124895

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 02257 WorkList ID : 168878 Department : Wet-Chemistry Date : 04-07-2023 16:24:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/06/2023	Water	O2221-01	pH <u>C</u>	Cool 4 deg C	PSEG03	M11	NWB-1904	04/06/2023	9040C
04/06/2023	Water	O2244-01	pH <u>A</u>	Cool 4 deg C	STER16	M11	MAIN-CELL	04/06/2023	9040C
04/06/2023	Water	O2244-02	pH <u>A</u>	Cool 4 deg C	STER16	M11	EX-CELL	04/06/2023	9040C
04/07/2023	Water	O2257-03	pH <u>Cy</u>	Cool 4 deg C	PSEG03	M11	ELI-WATER-0407	04/07/2023	9040C

Date/Time 04/07/23 16:00
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD CSM

Date/Time 04/07/23 18:00
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD CSM