



PERCENT SOLID

Supervisor: Sohil

Analyst: JIGNESH

Date: 4/10/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 14:25

In Date: 04/09/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:14

Out Date: 04/10/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB124901

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2246-01	EYG56	1	1.12	8.46	9.58	8.23	84.0	
O2246-02	EYJD9	2	1.11	8.77	9.88	8.45	83.7	
O2246-03	EYJE0	3	1.15	8.54	9.69	8.4	84.9	
O2246-04	EYJE1	4	1.15	8.40	9.55	8.26	84.6	
O2246-05	EYJE2	5	1.19	8.44	9.63	7.54	75.2	
O2246-06	EYJE3	6	1.1	8.49	9.59	8.03	81.6	
O2246-07	EYJE4	7	1.16	8.40	9.56	8.19	83.7	
O2246-08	EYJE5	8	1.19	8.67	9.86	8.37	82.8	
O2246-09	EYJE6	9	1.18	8.48	9.66	8.16	82.3	
O2246-10	EYJE7	10	1.11	8.55	9.66	8.00	80.6	
O2246-11	EYJE8	11	1.19	8.58	9.77	8.5	85.2	
O2246-12	EYJE9	12	1.18	8.56	9.74	8.57	86.3	
O2246-13	EYJF0	13	1.19	8.32	9.51	8.66	89.8	
O2246-14	EYJF1	14	1.18	8.66	9.84	8.71	87.0	
O2246-15	EYJF2	15	1.16	8.68	9.84	7.94	78.1	
O2246-16	EYJF3	16	1.18	8.59	9.77	8.22	82.0	
O2246-17	EYJF4	17	1.13	8.48	9.61	8.08	82.0	
O2246-18	EYJF5	18	1.15	8.82	9.97	8.61	84.6	
O2246-19	EYJF5MS	19	1.15	8.82	9.97	8.61	84.6	
O2246-20	EYJF5MSD	20	1.15	8.82	9.97	8.61	84.6	
O2246-21	EYJF8	21	1.17	8.53	9.7	8.42	85.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$