

LB124910


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49421



A= 0.0000e+000

B= 1.9935e-004

C= 1.9706e-002

Rho= 0.9996288

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.039	0.039	96	0.000	96	0				-11
0.2	0.200	0.178	-0.022	796	0.0 %	796					-5
2.5	2.500	2.384	-0.116	11862	0.0 %	11862					1
5.0	5.000	5.074	0.074	25353	0.0 %	25353					2
7.5	7.500	7.653	0.153	38291	0.0 %	38291					-1
10.0	10.000	9.872	-0.128	49421	0.0 %	49421					

LB124910 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Apr 2023 13:52:45

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	96	-	0.0000	7470A	PPB	10 Apr 2023 13:55:48	S	Std
0.2 - 1	50.2	796	-	0.2000	7470A	PPB	10 Apr 2023 13:58:07	S	Std
2.5 - 1	52.5	11862	-	2.5000	7470A	PPB	10 Apr 2023 14:00:24	S	Std
5.0 - 1	55	25353	-	5.0000	7470A	PPB	10 Apr 2023 14:02:42	S	Std
7.5 - 1	57.5	38291	-	7.5000	7470A	PPB	10 Apr 2023 14:04:58	S	Std
10.0 - 1	510	49421	-	10.0000	7470A	PPB	10 Apr 2023 14:07:15	S	Std
ICV12 - 1	ICV12	21155	4.2369	-7470A	PPB	PPB	10 Apr 2023 14:09:59	U	SMPL
ICB12 - 1	ICB12	-115	-0.0032	-7470A	PPB	PPB	10 Apr 2023 14:12:15	U	SMPL
CCV37 - 1	CCV37	25239	5.0510	-7470A	PPB	PPB	10 Apr 2023 14:14:31	U	SMPL
CCB37 - 1	CCB37	-108	-0.0018	-7470A	PPB	PPB	10 Apr 2023 14:16:46	U	SMPL
CRA - 1	CRA	810	0.1812	-7470A	PPB	PPB	10 Apr 2023 14:19:03	U	SMPL
HighStd - 1	HighStd	49894	9.9660	-7470A	PPB	PPB	10 Apr 2023 14:21:18	U	SMPL
ChkStd - 1	ChkStd	35986	7.1934	-7470A	PPB	PPB	10 Apr 2023 14:23:34	U	SMPL
PB152019BL - 1	PBW	-194	-0.0190	-7470A	PPB	PPB	10 Apr 2023 14:25:52	U	SMPL
PB152019BS - 1	LCSW	17865	3.5811	-7470A	PPB	PPB	10 Apr 2023 14:28:11	U	SMPL
O1952-07 - 1	PT-HG-WP	82837	16.5331	-7470A	PPB	PPB	10 Apr 2023 14:30:27	U	SMPL
O2257-03 - 1	ELI-WATER-0407	3954	0.8079	-7470A	PPB	PPB	10 Apr 2023 14:37:21	U	SMPL
O2257-03DUP - 1	ELI-WATER-0407DUP	3785	0.7742	-7470A	PPB	PPB	10 Apr 2023 14:39:40	U	SMPL
O2257-03MS - 1	ELI-WATER-0407MS	15139	3.0376	-7470A	PPB	PPB	10 Apr 2023 14:41:57	U	SMPL
O2257-03MSD - 1	ELI-WATER-0407MSD	15592	3.1279	-7470A	PPB	PPB	10 Apr 2023 14:44:14	U	SMPL
CCV38 - 1	CCV38	24682	4.9400	-7470A	PPB	PPB	10 Apr 2023 14:46:31	U	SMPL
CCB38 - 1	CCB38	55	0.0307	-7470A	PPB	PPB	10 Apr 2023 14:48:49	U	SMPL
O2257-03LX5 - 1		832	0.1856	-7470A	PPB	PPB	10 Apr 2023 15:02:51	U	SMPL
O2257-03A - 1		23180	4.6406	-7470A	PPB	PPB	10 Apr 2023 15:05:06	U	SMPL
O1952-07DLX5 - 1	PT-HG-WP	15444	3.0984	-7470A	PPB	PPB	10 Apr 2023 15:09:41	U	SMPL
CCV39 - 1	CCV39	25243	5.0518	-7470A	PPB	PPB	10 Apr 2023 15:14:19	U	SMPL
CCB39 - 1	CCB39	-63	0.0071	-7470A	PPB	PPB	10 Apr 2023 15:16:37	U	SMPL