



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124911

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/10/2023	14:10
2	CAL2	1	Water	NA	NA	20.7	7.01	04/10/2023	14:11
3	CAL3	1	Water	NA	NA	20.7	10.03	04/10/2023	14:12
4	ICV	1	Water	NA	NA	20.8	6.98	04/10/2023	14:15
5	CCV1	1	Water	NA	NA	20.8	2.01	04/10/2023	14:20
6	O2267-03	1	Water	NA	NA	20.5	6.45	04/10/2023	14:25
7	O2267-03DUP	1	Water	NA	NA	20.6	6.46	04/10/2023	14:26
8	O2268-01	1	Water	NA	NA	20.7	9.20	04/10/2023	14:35
9	CCV2	1	Water	NA	NA	20.8	12.03	04/10/2023	14:38

WB 124911

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 02267

WorkList ID : 168933

Department : Wet-Chemistry

Date : 04-10-2023 13:44:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/10/2023	Water	O2267-03	pH	Cool 4 deg C	PSEG03	M11	33988	04/10/2023	9040C
04/10/2023	Water	O2268-01	pH	Cool 4 deg C	PSEG03	M11	50002-TOTE	04/10/2023	9040C

Date/Time 04.10.23 13:50

Raw Sample Received by: 7D (QC)

Raw Sample Relinquished by: CP SM

Date/Time 04.10.23 15:00

Raw Sample Received by: CP SM

Raw Sample Relinquished by: 7D (QC)