

L3124912


SFAM01.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

52415



A= 0.0000e+000

B= 1.9124e-004 slope

C= 4.6740e-002 y-intercept

Rho= 0.9998746

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o D
0.00	0.000	0.066	0.066	102	0.000	102					11
0.05	0.050					438	0				12
0.20	0.200	0.224	0.024	926	0.0 %	926					-2
2.50	2.500	2.448	-0.052	12556	0.0 %	12556					-2
5.00	5.000	4.915	-0.085	25454	0.0 %	25454					0
7.50	7.500	7.477	-0.023	38852	0.0 %	38852					1
10.0	10.000	10.071	0.071	52415	0.0 %	52415					

LB124912

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	102	-		0 SFAM01.1	PPB	4/10/2023 15:40	Std
	0.2 S01	926	-		0.2 SFAM01.1	PPB	4/10/2023 15:43	Std
	2.5 S02	12556	-		2.5 SFAM01.1	PPB	4/10/2023 15:45	Std
	5 S03	25454	-		5 SFAM01.1	PPB	4/10/2023 15:49	Std
	7.5 S04	38852	-		7.5 SFAM01.1	PPB	4/10/2023 15:51	Std
	10 S05	52415	-		10 SFAM01.1	PPB	4/10/2023 15:53	Std
ICV013	ICV013	21358	4.1312 -		SFAM01.1	PPB	4/10/2023 15:56	SMPL
ICB013	ICB013	9	0.0485 -		SFAM01.1	PPB	4/10/2023 15:59	SMPL
CCV040	CCV40	25505	4.9243 -		SFAM01.1	PPB	4/10/2023 16:01	SMPL
CCB040	CCB40	28	0.0521 -		SFAM01.1	PPB	4/10/2023 16:03	SMPL
PB152022BL	PBS022	26	0.0517 -		SFAM01.1	PPB	4/10/2023 16:05	SMPL
O2116-01	MCB988	882	0.2154 -		SFAM01.1	PPB	4/10/2023 16:08	SMPL
O2116-02	MCB9C0	825	0.2045 -		SFAM01.1	PPB	4/10/2023 16:10	SMPL
O2116-03	MCB9C1	2381	0.5021 -		SFAM01.1	PPB	4/10/2023 16:12	SMPL
O2116-04	MCB9C2	1744	0.3803 -		SFAM01.1	PPB	4/10/2023 16:15	SMPL
O2116-05	MCB9C7	1541	0.3414 -		SFAM01.1	PPB	4/10/2023 16:17	SMPL
O2116-06	MCB9C7D	1436	0.3214 -		SFAM01.1	PPB	4/10/2023 16:19	SMPL
O2116-07	MCB9C7S	14681	2.8543 -		SFAM01.1	PPB	4/10/2023 16:21	SMPL
PB152023BL	POBS023	-15	0.0439 -		SFAM01.1	PPB	4/10/2023 16:24	SMPL
O2181-04	DCBJ5	548853	HIGH -		SFAM01.1	PPB	4/10/2023 16:28	
O2181-05	MDC8G0	81	0.0622 -		SFAM01.1	PPB	4/10/2023 16:30	SMPL
O2181-06	MDC8G1	1130	0.2628 -		SFAM01.1	PPB	4/10/2023 16:32	SMPL
O2181-07	MDC8G2	559	0.1536 -		SFAM01.1	PPB	4/10/2023 16:35	SMPL
O2181-08	MDC8G2D	637	0.1686 -		SFAM01.1	PPB	4/10/2023 16:37	SMPL
O2181-09	MDC8G2S	13237	2.5782 -		SFAM01.1	PPB	4/10/2023 16:39	SMPL
O2181-10	MDC8G3	1410	0.3164 -		SFAM01.1	PPB	4/10/2023 16:41	SMPL
O2181-11	MDC8G4	848	0.2089 -		SFAM01.1	PPB	4/10/2023 16:44	SMPL
O2181-12	MDC8G5	1004	0.2387 -		SFAM01.1	PPB	4/10/2023 16:46	SMPL
O2181-13	MDC8G6	986	0.2353 -		SFAM01.1	PPB	4/10/2023 16:48	SMPL
CCV041	CCV041	25560	4.9348 -		SFAM01.1	PPB	4/10/2023 16:51	SMPL
CCB041	CCB041	13	0.0492 -		SFAM01.1	PPB	4/10/2023 16:53	SMPL
O2181-14	MDC8G7	784	0.1967 -		SFAM01.1	PPB	4/10/2023 16:55	SMPL
O2181-15	MDC8H1	644	0.1699 -		SFAM01.1	PPB	4/10/2023 16:57	SMPL
O2181-03	DCBJ4	13100	2.552 -		SFAM01.1	PPB	4/10/2023 17:02	SMPL
O2181-04DLX10	DCBJ5	86081	16.5089 -		SFAM01.1	PPB	4/10/2023 17:08	SMPL
O2181-04DL2X50	DCBJ5	16997	3.2973 -		SFAM01.1	PPB	4/10/2023 17:18	SMPL
CCV042	CCV042	24799	4.7893 -		SFAM01.1	PPB	4/10/2023 17:22	SMPL
CCB042	CCB042	-137	0.0205 -		SFAM01.1	PPB	4/10/2023 17:24	SMPL