



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124913

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/10/2023	13:10
2	CAL2	1	Water	NA	NA	20.7	6.98	04/10/2023	13:11
3	CAL3	1	Water	NA	NA	20.7	10.03	04/10/2023	13:11
4	ICV	1	Water	NA	NA	20.8	6.99	04/10/2023	13:15
5	CCV1	1	Water	NA	NA	20.8	2.01	04/10/2023	13:22
6	O2257-01	1	Solid	20.02	20	20.6	7.66	04/10/2023	13:25
7	O2257-01DUP	1	Solid	20.01	20	20.7	7.67	04/10/2023	13:26
8	O2262-04	1	Solid	20.03	20	22.6	8.75	04/10/2023	13:30
9	O2262-08	1	Solid	20.02	20	22.5	9.24	04/10/2023	13:35
10	O2262-12	1	Solid	20.05	20	22.1	8.92	04/10/2023	13:40
11	CCV2	1	Water	NA	NA	20.8	12.02	04/10/2023	13:44

124913

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 02262 WorkList ID : 168920 Department : Wet-Chemistry Date : 04-10-2023 09:25:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/07/2023	Solid	O2257-01	Corrosivity	Cool 4 deg C	PSEG03	M11	ELI-SOLID-0407	04/07/2023	9045D
04/07/2023	Solid	O2262-04	Corrosivity	Cool 4 deg C	RUTW01	M11	JPP-2-040723	04/07/2023	9045D
04/07/2023	Solid	O2262-08	Corrosivity	Cool 4 deg C	RUTW01	M11	JPP-1-040723	04/07/2023	9045D
04/07/2023	Solid	O2262-12	Corrosivity	Cool 4 deg C	RUTW01	M11	JPP-3A-040723	04/07/2023	9045D

Date/Time 04.10.23 13:00
 Raw Sample Received by: JPC
 Raw Sample Relinquished by: JPC

Date/Time 04.10.23 16:30
 Raw Sample Received by: JPC
 Raw Sample Relinquished by: JPC