



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB124914
Thermometer ID: Flashpoint

Reviewed By: Rubina
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 776.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/10/2023	13:55
2	O2257-03		1	14	100.00	>212.0	>212.0	04/10/2023	14:25
3	O2257-03DUP		1	13	100.00	>212.0	>212.0	04/10/2023	14:55
4	O2267-03		1	14	100.00	>212.0	>212.0	04/10/2023	15:25
5	O2268-01		1	12	100.00	>212.0	>212.0	04/10/2023	15:55

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

661249124

WorkList Name : fp-4-10

WorkList ID : 168905

Department : Wet-Chemistry

Date : 04-10-2023 07:54:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/14/2023	Water	O2257-03	Flash Point	Cool 4 deg C	PSEG03	M11	ELI-WATER-0407	04/07/2023	1010B
04/17/2023	Water	O2267-03	Flash Point	Cool 4 deg C	PSEG03	M11	33988	04/10/2023	1010B
04/13/2023	Water	O2268-01	Flash Point	Cool 4 deg C	PSEG03	M11	50002-TOTE	04/10/2023	1010B

Date/Time 04/10/2023 13:45
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: AD(CWC)

Date/Time 04/10/2023 16:30
 Raw Sample Received by: AD(CWC)
 Raw Sample Relinquished by: RM CWL