

LB124920


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49932



A= 0.0000e+000

B= 2.0121e-004

C= 3.5142e-003

Rho= 0.9995657

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10 D
0.0	0.000	0.082	0.082	391	0.000	391	0				
0.2	0.200	0.207	0.007	1010	0.0 %	1010					4
2.5	2.500	2.524	0.024	12528	0.0 %	12528					1
5.0	5.000	4.764	-0.236	23659	0.0 %	23659					5
7.5	7.500	7.573	0.073	37619	0.0 %	37619					1
10.0	10.000	10.050	0.050	49932	0.0 %	49932					1

LB124920 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Apr 2023 10:51:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	391	-	0.0000	7470A	PPB	11 Apr 2023 10:57:51	S	Std
0.2 - 1	50.2	1010	-	0.2000	7470A	PPB	11 Apr 2023 11:00:09	S	Std
2.5 - 1	52.5	12528	-	2.5000	7470A	PPB	11 Apr 2023 11:02:26	S	Std
5.0 - 1	55	23659	-	5.0000	7470A	PPB	11 Apr 2023 11:04:43	S	Std
7.5 - 1	57.5	37619	-	7.5000	7470A	PPB	11 Apr 2023 11:07:00	S	Std
10.0 - 1	510	49932	-	10.0000	7470A	PPB	11 Apr 2023 11:11:53	S	Std
ICV14 - 1	ICV14	21557	4.3409	-	7470A	PPB	11 Apr 2023 11:14:47	U	SMPL
ICB14 - 1	ICB14	-75	-0.0116	-	7470A	PPB	11 Apr 2023 11:17:02	U	SMPL
CCV43 - 1	CCV43	24935	5.0206	-	7470A	PPB	11 Apr 2023 11:19:19	U	SMPL
CCB43 - 1	CCB43	-135	-0.0236	-	7470A	PPB	11 Apr 2023 11:21:34	U	SMPL
CRA - 1	CRA	829	0.1703	-	7470A	PPB	11 Apr 2023 11:23:51	U	SMPL
HighStd - 1	HighStd	49408	9.9448	-	7470A	PPB	11 Apr 2023 11:26:06	U	SMPL
ChkStd - 1	ChkStd	36020	7.2510	-	7470A	PPB	11 Apr 2023 11:28:21	U	SMPL
PB152024BL - 1	PBW	-199	-0.0365	-	7470A	PPB	11 Apr 2023 11:30:40	U	SMPL
PB152024BS - 1	LCSW	19060	3.8385	-	7470A	PPB	11 Apr 2023 11:32:58	U	SMPL
O2224-02 - 1	OU3-FC-BACKFILL-01	4541	0.9172	-	7470A	PPB	11 Apr 2023 11:35:15	U	SMPL
O2224-04 - 1	OU3-FC-BACKFILL-02	919	0.1884	-	7470A	PPB	11 Apr 2023 11:37:32	U	SMPL
O2224-06 - 1	OU3-FC-BACKFILL-03	486	0.1013	-	7470A	PPB	11 Apr 2023 11:39:51	U	SMPL
O2224-08 - 1	OU3-FC-BACKFILL-04	407	0.0854	-	7470A	PPB	11 Apr 2023 11:42:10	U	SMPL
O2224-10 - 1	OU3-FC-BACKFILL-05	382	0.0804	-	7470A	PPB	11 Apr 2023 11:44:25	U	SMPL
CCV44 - 1	CCV44	24904	5.0144	-	7470A	PPB	11 Apr 2023 11:46:41	U	SMPL
CCB44 - 1	CCB44	35	0.0106	-	7470A	PPB	11 Apr 2023 11:48:56	U	SMPL
O2224-12 - 1	OU3-FC-BACKFILL-06	610	0.1263	-	7470A	PPB	11 Apr 2023 11:51:11	U	SMPL
O2224-14 - 1	OU3-FC-BACKFILL-07	278	0.0594	-	7470A	PPB	11 Apr 2023 11:53:27	U	SMPL
O2224-16 - 1	OU3-FC-BACKFILL-08	192	0.0421	-	7470A	PPB	11 Apr 2023 11:55:42	U	SMPL
O2224-18 - 1	OU3-FC-BACKFILL-09	313	0.0665	-	7470A	PPB	11 Apr 2023 11:57:58	U	SMPL
O2224-20 - 1	OU3-FC-BACKFILL-10	337	0.0713	-	7470A	PPB	11 Apr 2023 12:00:14	U	SMPL
O2224-22 - 1	OU3-FC-BACKFILL-11	1444	0.2941	-	7470A	PPB	11 Apr 2023 12:02:31	U	SMPL
O2224-24 - 1	OU3-FC-BACKFILL-12	259	0.0556	-	7470A	PPB	11 Apr 2023 12:04:47	U	SMPL
O2224-26 - 1	OU3-FC-BACKFILL-13	174	0.0385	-	7470A	PPB	11 Apr 2023 12:07:04	U	SMPL
O2224-28 - 1	OU3-FC-BACKFILL-14	159	0.0355	-	7470A	PPB	11 Apr 2023 12:09:21	U	SMPL
O2224-30 - 1	OU3-FC-BACKFILL-15	199	0.0436	-	7470A	PPB	11 Apr 2023 12:11:37	U	SMPL
CCV45 - 1	CCV45	24622	4.9576	-	7470A	PPB	11 Apr 2023 12:13:53	U	SMPL
CCB45 - 1	CCB45	38	0.0112	-	7470A	PPB	11 Apr 2023 12:16:08	U	SMPL
O2224-30DUP - 1	OU3-FC-BACKFILL-15DUP	466	0.0973	-	7470A	PPB	11 Apr 2023 12:18:24	U	SMPL
O2224-30MS - 1	OU3-FC-BACKFILL-15MS	17966	3.6184	-	7470A	PPB	11 Apr 2023 12:20:40	U	SMPL
O2224-30MSD - 1	OU3-FC-BACKFILL-15MSD	19576	3.9424	-	7470A	PPB	11 Apr 2023 12:22:55	U	SMPL
PB152024TB - 1	PB151972TB	-5	0.0025	-	7470A	PPB	11 Apr 2023 12:25:12	U	SMPL
O2224-30LX5 - 1		37	0.0110	-	7470A	PPB	11 Apr 2023 12:27:28	U	SMPL
O2224-30A - 1		21520	4.3335	-	7470A	PPB	11 Apr 2023 12:29:45	U	SMPL
CCV46 - 1	CCV46	24652	4.9637	-	7470A	PPB	11 Apr 2023 12:32:01	U	SMPL
CCB46 - 1	CCB46	-1	0.0033	-	7470A	PPB	11 Apr 2023 12:34:19	U	SMPL