

LB124923


SFAM01.1

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

48092



A= 0.0000e+000

B= 2.0578e-004 slope

C= 3.6407e-002 $y = mx + c$

Rho= 0.9996826

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/D
0.00	0.000	0.071	0.071	170	0.000	170					1
0.05	0.050					438	0				1
0.20	0.200	0.205	0.005	818	0.0 %	818					3
2.50	2.500	2.398	-0.102	11476	0.0 %	11476					4
5.00	5.000	4.930	-0.070	23779	0.0 %	23779					1
7.50	7.500	7.664	0.164	37066	0.0 %	37066					2
10.0	10.000	9.933	-0.067	48092	0.0 %	48092					1

LB124923

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	170	-		0 SFAM01.1	PPB	4/11/2023 16:03	Std
	0.2 S01	818	-		0.2 SFAM01.1	PPB	4/11/2023 16:05	Std
	2.5 S02	11476	-		2.5 SFAM01.1	PPB	4/11/2023 16:07	Std
	5 S03	23779	-		5 SFAM01.1	PPB	4/11/2023 16:10	Std
	7.5 S04	37066	-		7.5 SFAM01.1	PPB	4/11/2023 16:12	Std
	10 S05	48092	-		10 SFAM01.1	PPB	4/11/2023 16:14	Std
ICV015	ICV015	20630	4.2816 -		SFAM01.1	PPB	4/11/2023 16:17	SMPL
ICB015	ICB015	-6	0.0352 -		SFAM01.1	PPB	4/11/2023 16:20	SMPL
CCV047	CCV047	23543	4.881 -		SFAM01.1	PPB	4/11/2023 16:22	SMPL
CCB047	CCB047	11	0.0387 -		SFAM01.1	PPB	4/11/2023 16:24	SMPL
PB152025BL	PBW025	39	0.0444 -		SFAM01.1	PPB	4/11/2023 16:26	SMPL
PB152025TB	LEB025	119	0.0609 -		SFAM01.1	PPB	4/11/2023 16:29	SMPL
O2227-01	MCOLF7	60	0.0488 -		SFAM01.1	PPB	4/11/2023 16:31	SMPL
O2227-02	MCOLF8	83	0.0535 -		SFAM01.1	PPB	4/11/2023 16:33	SMPL
O2227-03	MCOLF8D	122	0.0615 -		SFAM01.1	PPB	4/11/2023 16:35	SMPL
O2227-04	MCOLF8S	4651	0.9935 -		SFAM01.1	PPB	4/11/2023 16:38	SMPL
CCV048	CCV048	23123	4.7946 -		SFAM01.1	PPB	4/11/2023 16:40	SMPL
CCB048	CCB048	27	0.042 -		SFAM01.1	PPB	4/11/2023 16:42	SMPL