



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124927

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	04/11/2023	16:20
2	CAL2	1	Water	NA	NA	20.7	7.00	04/11/2023	16:21
3	CAL3	1	Water	NA	NA	20.7	10.03	04/11/2023	16:22
4	ICV	1	Water	NA	NA	20.7	6.98	04/11/2023	16:25
5	CCV1	1	Water	NA	NA	20.7	2.02	04/11/2023	17:05
6	O2277-03	1	Water	NA	NA	22.4	8.12	04/11/2023	17:14
7	O2288-02	1	Water	NA	NA	20.8	6.22	04/11/2023	17:20
8	O2288-02DUP	1	Water	NA	NA	20.9	6.23	04/11/2023	17:21
9	CCV2	1	Water	NA	NA	20.8	12.02	04/11/2023	17:25

183 124927

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY O2288

WorkList ID : 168959

Department : Wet-Chemistry

Date : 04-11-2023 16:03:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/11/2023	Water	O2277-03	Corrosivity	Cool 4 deg C	PSEG03	M11	SOM-23-00002	04/11/2023	9040C
04/11/2023	Water	O2288-02	Corrosivity	Cool 4 deg C	SEAC01	M11	GROUND-WATER	04/11/2023	9040C

Date/Time 04.11.23 16:03
 Raw Sample Received by: 781001
 Raw Sample Relinquished by: 121001

Date/Time 04.11.23 18:00
 Raw Sample Received by: 121001
 Raw Sample Relinquished by: 781001