



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124928

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	04/11/2023	16:20
2	CAL2	1	Water	NA	NA	20.7	7.00	04/11/2023	16:21
3	CAL3	1	Water	NA	NA	20.7	10.03	04/11/2023	16:22
4	ICV	1	Water	NA	NA	20.7	6.98	04/11/2023	16:25
5	CCV1	1	Water	NA	NA	20.6	2.01	04/11/2023	17:30
6	O2278-01	1	Water	NA	NA	20.6	4.33	04/11/2023	17:33
7	O2286-04	1	Water	NA	NA	21.5	5.51	04/11/2023	17:38
8	O2291-01	1	Water	NA	NA	22.1	8.41	04/11/2023	17:42
9	O2291-02	1	Water	NA	NA	21.8	8.82	04/11/2023	17:45
10	O2291-02DUP	1	Water	NA	NA	21.9	8.83	04/11/2023	17:46
11	CCV2	1	Water	NA	NA	20.7	12.02	04/11/2023	17:49

WORKLIST(Hardcopy Internal Chain)

VB 24928
VB 24928

WorkList Name : ph o2291

WorkList ID : 168958

Department : Wet-Chemistry

Date : 04-11-2023 16:02:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/05/2023	Water	O2278-01	pH	Cool 4 deg C	STER11	M11	POOL-WATER	04/05/2023	9040C
04/11/2023	Water	O2286-04	pH	Cool 4 deg C	PSEG03	M11	328	04/11/2023	9040C
04/11/2023	Water	O2291-01	pH	Cool 4 deg C	PSEG03	M11	23330	04/11/2023	9040C
04/11/2023	Water	O2291-02	pH	Cool 4 deg C	PSEG03	M11	2344	04/11/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

04-11-23 16:07
JP (QPC)
12/24

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

04-11-23 18:00
12/24
JP (QPC)