

23124937


SFAM01.1

INSTRUMENTAL DISCV1

Linear

μ Abs.:

43919



A= 0.0000e+000

B= 2.2426e-004 slope

C= 4.0917e-002 y-intercept

Rho= 0.9994133

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.00	0.000	0.071	0.071	135	0.000	135					11
0.05	0.050					438	0				2
0.20	0.200	0.203	0.003	724	0.0 %	724					4
2.50	2.500	2.410	-0.090	10562	0.0 %	10562					1
5.00	5.000	4.886	-0.114	21606	0.0 %	21606					3
7.50	7.500	7.740	0.240	34331	0.0 %	34331					1
10.0	10.000	9.890	-0.110	43919	0.0 %	43919					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	135	-		0 SFAM01.1	PPB	4/12/2023 13:37	Std
	0.2 S01	724	-		0.2 SFAM01.1	PPB	4/12/2023 13:39	Std
	2.5 S02	10562	-		2.5 SFAM01.1	PPB	4/12/2023 13:42	Std
	5 S03	21606	-		5 SFAM01.1	PPB	4/12/2023 13:44	Std
	7.5 S04	34331	-		7.5 SFAM01.1	PPB	4/12/2023 13:46	Std
	10 S05	43919	-		10 SFAM01.1	PPB	4/12/2023 13:48	Std
ICV017	ICV017	18099	4.0997 -		SFAM01.1	PPB	4/12/2023 13:53	SMPL
ICB017	ICB017	-8	0.0391 -		SFAM01.1	PPB	4/12/2023 13:55	SMPL
CCV054	CCV054	21915	4.9555 -		SFAM01.1	PPB	4/12/2023 13:57	SMPL
CCB054	CCB054	-24	0.0355 -		SFAM01.1	PPB	4/12/2023 13:59	SMPL
PB152080BL	PBW080	67	0.0559 -		SFAM01.1	PPB	4/12/2023 14:02	SMPL
O2169-01	EYG74	76	0.058 -		SFAM01.1	PPB	4/12/2023 14:04	SMPL
O2169-02	EYG74D	203	0.0864 -		SFAM01.1	PPB	4/12/2023 14:06	SMPL
O2169-03	EYG74S	4281	1.001 -		SFAM01.1	PPB	4/12/2023 14:08	SMPL
O2169-04	EYG75	106	0.0647 -		SFAM01.1	PPB	4/12/2023 14:11	SMPL
O2169-05	EYG76	162	0.0772 -		SFAM01.1	PPB	4/12/2023 14:13	SMPL
O2169-06	EYG80	28	0.0472 -		SFAM01.1	PPB	4/12/2023 14:15	SMPL
PB152081BL	PBW081	47	0.0515 -		SFAM01.1	PPB	4/12/2023 14:17	SMPL
O2181-01	DCBJ1	22060	4.988 -		SFAM01.1	PPB	4/12/2023 14:22	SMPL
O2181-16	MDC8H2	786	0.2172 -		SFAM01.1	PPB	4/12/2023 14:31	SMPL
O2181-17	MDC8H2D	684	0.1943 -		SFAM01.1	PPB	4/12/2023 14:36	SMPL
O2181-18	MDC8H2S	4904	1.1407 -		SFAM01.1	PPB	4/12/2023 14:38	SMPL
O2181-19	MDC8H3	229	0.0923 -		SFAM01.1	PPB	4/12/2023 14:41	SMPL
O2181-20	MDC8H4	469	0.1461 -		SFAM01.1	PPB	4/12/2023 14:43	SMPL
CCV055	CCV055	21329	4.8241 -		SFAM01.1	PPB	4/12/2023 14:45	SMPL
CCB055	CCB055	71	0.0568 -		SFAM01.1	PPB	4/12/2023 14:48	SMPL