



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124942

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	04/12/2023	10:10
2	O2277-03		1	14	100.00	>212.0	>212.0	04/12/2023	10:40
3	O2277-03DUP		1	12	100.00	>212.0	>212.0	04/12/2023	11:10
4	O2286-04		1	14	100.00	>212.0	>212.0	04/12/2023	11:40
5	O2288-02		1	14	100.00	>212.0	>212.0	04/12/2023	12:10
6	O2291-01		1	12	100.00	>212.0	>212.0	04/12/2023	12:40
7	O2291-02		1	12	100.00	>212.0	>212.0	04/12/2023	13:10
8	O2293-01		1	14	100.00	>212.0	>212.0	04/12/2023	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB12WQA42

WorkList Name : fp-4-11

WorkList ID : 168941

Department : Wet-Chemistry

Date : 04-11-2023 07:52:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/24/2023	Water	O2288-02H	Flash Point	Cool 4 deg C	SEAC01	M11	GROUND-WATER	04/11/2023	1010B
04/13/2023	Water	O2293-01A	Flash Point	Cool 4 deg C	JACO05	M11	IDW-LIQ-FRAC-041123	04/11/2023	1010B
04/18/2023	Water	O2277-03B	Flash Point	Cool 4 deg C	PSEG03	M11	SOM-23-00002	04/11/2023	1010B
04/18/2023	Water	O2286-04B	Flash Point	Cool 4 deg C	PSEG03	M11	328	04/11/2023	1010B
04/18/2023	Water	O2291-01C	Flash Point	Cool 4 deg C	PSEG03	M11	23330	04/11/2023	1010B
04/18/2023	Water	O2291-02D	Flash Point	Cool 4 deg C	PSEG03	M11	2344	04/11/2023	1010B

Date/Time 04/12/2023 10:00
 Raw Sample Received by: RM Cwe2 (201)
 Raw Sample Relinquished by: RM Cwe2

Date/Time 04/12/2023 10:00
 Raw Sample Received by: RM Cwe2 (201)
 Raw Sample Relinquished by: RM Cwe2