



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124943

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	04/12/2023	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	04/12/2023	16:26
3	CAL3	1	Water	NA	NA	20.7	10.02	04/12/2023	16:29
4	ICV	1	Water	NA	NA	20.8	6.98	04/12/2023	16:30
5	CCV1	1	Water	NA	NA	20.8	2.02	04/12/2023	17:10
6	O2298-01	1	Water	NA	NA	20.7	4.93	04/12/2023	17:15
7	O2298-02	1	Water	NA	NA	20.9	4.78	04/12/2023	17:22
8	O2298-03	1	Water	NA	NA	21.0	6.31	04/12/2023	17:28
9	O2303-01	1	Water	NA	NA	20.7	7.40	04/12/2023	17:30
10	O2319-01	1	Water	NA	NA	22.5	6.51	04/12/2023	17:33
11	O2319-01DUP	1	Water	NA	NA	22.6	6.52	04/12/2023	17:34
12	CCV2	1	Water	NA	NA	20.8	12.03	04/12/2023	17:36

VB 124943

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 02319 WorkList ID : 169001 Department : Wet-Chemistry Date : 04-12-2023 15:02:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/05/2023	Water	O2298-01	pH	Cool 4 deg C	STER09	M11	CELL-A-POOL-WATER	04/05/2023	9040C
04/05/2023	Water	O2298-02	pH	Cool 4 deg C	STER09	M11	CELL-B-C-POOL-WATER	04/05/2023	9040C
04/05/2023	Water	O2298-03	pH	Cool 4 deg C	STER09	M11	CELL-D-POOL-WATER	04/05/2023	9040C
04/12/2023	Water	O2303-01	pH	Cool 4 deg C	PSEG04	M11	402	04/12/2023	9040C
04/11/2023	Water	O2319-01	pH	Cool 4 deg C	ENTA05	M11	E9074-PVSC-002	04/11/2023	9040C

Date/Time 04-12-23 16:20

Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 04.12.23 18:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]