



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124944

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	04/12/2023	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	04/12/2023	16:26
3	CAL3	1	Water	NA	NA	20.7	10.02	04/12/2023	16:29
4	ICV	1	Water	NA	NA	20.8	6.98	04/12/2023	16:30
5	CCV1	1	Water	NA	NA	20.7	2.01	04/12/2023	16:32
6	O2292-02	1	Solid	20.02	20	22.3	9.12	04/12/2023	16:40
7	O2292-04	1	Solid	20.03	20	21.7	9.00	04/12/2023	16:41
8	O2292-06	1	Solid	20.05	20	21.3	7.10	04/12/2023	16:45
9	O2292-08	1	Solid	20.04	20	22.0	7.14	04/12/2023	16:48
10	O2292-10	1	Solid	20.01	20	21.0	6.49	04/12/2023	16:50
11	O2305-02	1	Solid	20.03	20	24.7	8.42	04/12/2023	16:52
12	O2310-01	1	Solid	20.01	20	24.6	9.22	04/12/2023	17:00
13	O2310-02	1	Solid	20.01	20	24.5	8.26	04/12/2023	17:01
14	O2310-02DUP	1	Solid	20.02	20	24.6	8.27	04/12/2023	17:03
15	CCV2	1	Water	NA	NA	20.7	12.03	04/12/2023	17:05

VB 124944

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 02292 WorkList ID : 168989 Department : Wet-Chemistry Date : 04-12-2023 10:27:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/11/2023	Solid	O2292-02	Corrosivity	Cool 4 deg C	JACO05	M11	IDW-SO-VB3-041123	04/11/2023	9045D
04/11/2023	Solid	O2292-04	Corrosivity	Cool 4 deg C	JACO05	M11	IDW-SO-VB1-041123	04/11/2023	9045D
04/11/2023	Solid	O2292-06	Corrosivity	Cool 4 deg C	JACO05	M11	IDW-SO-G1-041123	04/11/2023	9045D
04/11/2023	Solid	O2292-08	Corrosivity	Cool 4 deg C	JACO05	M11	IDW-SO-G2-041123	04/11/2023	9045D
04/11/2023	Solid	O2292-10	Corrosivity	Cool 4 deg C	JACO05	M11	IDW-SO-G3-041123	04/11/2023	9045D
04/12/2023	Solid	O2305-02	Corrosivity	Cool 4 deg C	PSEG03	M11	VNJ-232	04/12/2023	9045D
04/12/2023	Solid	O2310-01	Corrosivity	Cool 4 deg C	PSEG03	M11	VNJ-228	04/12/2023	9045D
04/12/2023	Solid	O2310-02	Corrosivity	Cool 4 deg C	PSEG03	M11	RBR-200059	04/12/2023	9045D

Date/Time 04.12.23 16:20
 Raw Sample Received by: 10 (GOC)
 Raw Sample Relinquished by: [Signature]

Date/Time 04.12.23 18:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]