



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB124945

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	04/12/2023	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	04/12/2023	16:26
3	CAL3	1	Water	NA	NA	20.7	10.02	04/12/2023	16:29
4	ICV	1	Water	NA	NA	20.8	6.98	04/12/2023	16:30
5	CCV1	1	Water	NA	NA	20.7	2.01	04/12/2023	16:32
6	O2293-01	1	Water	NA	NA	20.2	9.05	04/12/2023	16:35
7	O2293-01DUP	1	Water	NA	NA	20.1	9.06	04/12/2023	16:36
8	CCV2	1	Water	NA	NA	20.8	12.02	04/12/2023	16:37

VB1249B5 50
VB124945

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 02293 W WorkList ID : 168990 Department : Wet-Chemistry Date : 04-12-2023 10:28:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/11/2023	Water	02293-01	Corrosivity -D	Cool 4 deg C	JACO05	M11	IDW-LIQ-FRAC-041123	04/11/2023	9040C

Date/Time 04-12-23 16:20
Raw Sample Received by: 70609C1
Raw Sample Relinquished by: [Signature]

Date/Time 04-12-23 18:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: 70609C1