

LB124957


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

6724E



A= 0.0000e+000

B= 1.4970e-004

C= 1.6620e-002

Rho= 0.9998028

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.028	0.028	76	0.000	76					7
0.2	0.200	0.215	0.015	1326	0.0 %	1326					2
2.5	2.500	2.554	0.054	16951	0.0 %	16951					3
5.0	5.000	4.860	-0.140	32354	0.0 %	32354					1
7.5	7.500	7.459	-0.041	49715	0.0 %	49715					1
10.0	10.000	10.084	0.084	67248	0.0 %	67248					

LB124957 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Apr 2023 14:28:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	76	-	0.0000	7471B	PPB	13 Apr 2023 14:28:48	S	Std
0.2 - 1	SO.2	1326	-	0.2000	7471B	PPB	13 Apr 2023 14:31:05	S	Std
2.5 - 1	SO.5	16951	-	2.5000	7471B	PPB	13 Apr 2023 14:33:23	S	Std
5.0 - 1	SB	32354	-	5.0000	7471B	PPB	13 Apr 2023 14:35:40	S	Std
7.5 - 1	SB.5	49715	-	7.5000	7471B	PPB	13 Apr 2023 14:37:56	S	Std
10.0 - 1	SIV	67248	-	10.0000	7471B	PPB	13 Apr 2023 14:45:29	S	Std
ICV20 - 1	ICV20	24517	3.6868	-7471B	PPB	13 Apr 2023 14:48:17	U	SMPL	
ICB20 - 1	ICB20	-83	0.0042	-7471B	PPB	13 Apr 2023 14:50:32	U	SMPL	
CCV61 - 1	CCV61	32553	4.8898	-7471B	PPB	13 Apr 2023 14:52:48	U	SMPL	
CCB61 - 1	CCB61	-74	0.0055	-7471B	PPB	13 Apr 2023 14:55:03	U	SMPL	
CRA - 1	CRA	1305	0.2120	-7471B	PPB	13 Apr 2023 14:57:20	U	SMPL	
HighStd - 1	HighStd	67988	10.1945	-7471B	PPB	13 Apr 2023 14:59:35	U	SMPL	
ChkStd - 1	ChkStd	47907	7.1883	-7471B	PPB	13 Apr 2023 15:01:51	U	SMPL	
PB152107BL - 1	PBS	-133	-0.0033	-7471B	PPB	13 Apr 2023 15:04:10	U	SMPL	
PB152107BS - 1	LCSS	22982	3.4570	-7471B	PPB	13 Apr 2023 15:06:28	U	SMPL	
O2283-01 - 1	20230323-COMP-1	67862	10.1756	-7471B	PPB	13 Apr 2023 15:08:45	U	SMPL	
O2283-02 - 1	20230326-COMP-4	29324	4.4064	-7471B	PPB	13 Apr 2023 15:11:03	U	SMPL	
O2283-03 - 1	20230327-COMP-5	4209	0.6467	-7471B	PPB	13 Apr 2023 15:13:23	U	SMPL	
O2283-04 - 1	20230328-COMP-6	1561	0.2503	-7471B	PPB	13 Apr 2023 15:15:41	U	SMPL	
O2283-05 - 1	20230329-COMP-7	12559	1.8967	-7471B	PPB	13 Apr 2023 15:17:57	U	SMPL	
CCV62 - 1	CCV62	32422	4.8702	-7471B	PPB	13 Apr 2023 15:20:12	U	SMPL	
CCB62 - 1	CCB62	-49	0.0093	-7471B	PPB	13 Apr 2023 15:22:28	U	SMPL	
O2283-06 - 1	20230334-COMP-12	18741	2.8222	-7471B	PPB	13 Apr 2023 15:24:45	U	SMPL	
O2305-01 - 1	VNJ-232	14712	2.2190	-7471B	PPB	13 Apr 2023 15:27:00	U	SMPL	
O2307-01 - 1	PL-02-041223	842	0.1427	-7471B	PPB	13 Apr 2023 15:29:17	U	SMPL	
O2307-01DUP - 1	PL-02-041223DUP	1177	0.1928	-7471B	PPB	13 Apr 2023 15:31:34	U	SMPL	
O2307-01MS - 1	PL-02-041223MS	28695	4.3123	-7471B	PPB	13 Apr 2023 15:33:50	U	SMPL	
O2308-01 - 1	EO-01-041223	2953	0.4587	-7471B	PPB	13 Apr 2023 15:36:24	U	SMPL	
O2320-01 - 1	E9074-BAYS-001	3220	0.4987	-7471B	PPB	13 Apr 2023 15:40:42	U	SMPL	
O2321-01 - 1	E9074-BAYC-001	849	0.1437	-7471B	PPB	13 Apr 2023 15:42:59	U	SMPL	
O2322-01 - 1	E9074-BAYC-002	376	0.0729	-7471B	PPB	13 Apr 2023 15:45:15	U	SMPL	
CCV63 - 1	CCV63	32739	4.9177	-7471B	PPB	13 Apr 2023 15:47:31	U	SMPL	
CCB63 - 1	CCB63	-28	0.0124	-7471B	PPB	13 Apr 2023 15:49:46	U	SMPL	
O2324-01 - 1	RT-3419	2665	0.4156	-7471B	PPB	13 Apr 2023 15:52:03	U	SMPL	
O2307-01MSD - 1	PL-02-041223MSD	27918	4.1960	-7471B	PPB	13 Apr 2023 15:54:33	U	SMPL	
O2307-01LX5 - 1		77	0.0281	-7471B	PPB	13 Apr 2023 15:57:09	U	SMPL	
O2307-01A - 1		27273	4.0994	-7471B	PPB	13 Apr 2023 15:59:24	U	SMPL	
O2283-01DLX2 - 1		32308	4.8531	-7471B	PPB	13 Apr 2023 16:01:39	U	SMPL	
CCV64 - 1	CCV64	32987	4.9548	-7471B	PPB	13 Apr 2023 16:03:57	U	SMPL	
CCB64 - 1	CCB64	-11	0.0150	-7471B	PPB	13 Apr 2023 16:06:15	U	SMPL	